Determination of Trace Amounts of Arsenic(III) and Arsenic(V) in Drinking Water and Arsenic(III) Vapor in Air by Graphite-Furnace Atomic Absorption Spectrophotometry Using 2,3-Dimercaptopropane-1-sulfonate as a Complexing Agent

Chuan-Jier HSIEH, Chia-Hsin YEN and Mao-Sung KUO[†]

Department of Environmental Science, Tunghai University, Taichung 407, Taiwan, Republic of China

As(III) (12.5 – 200 ng) in drinking water (25 ml) reacted with 2,3-dimercaptopropane-1-sulfonate (DMPS, 0.60 mg) in ammonium acetate buffer (2.0 mmol, pH 5.5) and formed a complex of As(III)-DMPS. The complex was selectively retained on two Sep-Pak C₁₈ cartridges in series, while As(V) could not be retained. Each cartridge was eluted with methanol (2.00 ml). After addition of Ni²⁺ (2.0 mg), a portion (20 µl) was introduced into a graphite cuvette and was atomized according to a temperature program. As(V) (up to 100 ng) in drinking water (25 ml) was pre-reduced to As(III) with L-cysteine. The concentration of As(V) was determined by subtracting As(III) from the final total concentration of As(III)+As(V). The method detection limits (3σ) for As(III) and As(V) were 0.11 µg/l and 0.15 µg/l, respectively; the calibration graphs were linear up to 8.00 µg/l for As(III). Accuracies of 97.6 – 99.3% were obtained for As(III) and As(V) in drinking water with a precision within 4.0%. The proposed method could be applied to the determination of As(III) vapor in air with a detection limit of 2.8 ng.

Keywords Arsenic(III), arsenic(V), drinking water, air, 2,3-dimercaptopropane-1-sulfonate, graphite-furnace atomic absorption spectrophotometry

Cumulative excess amounts of arsenic in our body may cause liver, lung, kidney, bladder or skin cancers, and chronic poisoning.¹⁻³ In the southwest coastal region of Taiwan (Tainan and Chia-I county)¹, a few residents had for more than 30 years been drinking deep groundwater containing high levels (600 – 700 µg/l) of arsenic compounds^{4,5}, which caused Blackfoot Disease. Based on this phenomenon, the USEPA and Taiwan government have considered decreasing the allowable contamination levels of total arsenic in drinking water to be 2 – 20 µg/l⁶ and 10 µg/l, respectively.

As(III) is considered to be more toxic than As(V), which is in turn more toxic than organic arsenics. The 8-h average permissible exposure limit for As₂O₃ in air was recommended not to exceed 0.01 mg/m³ of As by OSHA.⁷ Several methods commonly used for the determination of the amounts of arsenic species in water are flow-injection with graphite-furnace atomic absorption spectrometry (FI-GFAAS)⁸, or with hydride generation atomic absorption spectrometry (FI-HGAAS);^{9,10} chelating extraction and then HGAAS¹¹, or neutron activation analysis;¹² HPLC and then HGAAS⁴, or HG-ICPMS;¹³ and ion chromatography-HG-ICPMS.¹⁴ The As₂O₃ vapor in air determined by NIOSH⁷ was collected on a Na₂CO₃-impregnated cellulose ester membrane. After digestion and the addition

of Ni²⁺, an aliquot (25 µl) was analyzed by GFAAS. The detection limit was 0.06 µg per sample and the linear range was 0.67 – 32 µg/m³ as As for a 400-l sample.

2,3-Dimercaptopropane-1-sulfonate (DMPS) has been used as an antidote for mice¹⁵ and human¹⁶ poisoning with sodium arsenite, and Ni²⁺ has been commonly used as a chemical modifier in GFAAS^{5,17,18} to form a stable compound of NiAs.¹⁹ In this paper, we used DMPS as a complexing agent for As(III) in ammonium acetate buffer of pH 5.5. The complex of As(III)-DMPS was selectively retained on Sep-Pak C₁₈ cartridges and then concentrated in methanol; while As(V) could not be retained in this way. The As(V) was pre-reduced to As(III) with L-cysteine^{9,13,20,21} and the total amount of As(III)+As(V) was determined. The amount of As(V) was obtained by subtracting As(III) from [As(III)+As(V)]. Hence, trace amounts of As(III) (up to 200 ng) and As(V) (up to 100 ng) in drinking water (25 ml) could be accurately determined by GFAAS. The applicability for the determination of As(III) vapor in air was also attempted.

Experimental

Apparatus

A Hitachi Z-8000 graphite-furnace atomic absorption spectrophotometer, equipped with a Zeeman back-

[†] To whom correspondence should be addressed.

ground corrector, was used for an atomic absorption measurement of arsenic at 193.7 nm with a slit width of 1.3 nm. A hollow cathode lamp of As (Hitachi Co., Japan) was operated at 10 mA. Uncoated graphite tube cuvettes (No. 180-7400, Hitachi Co., Japan) were purchased. An air sample pump (224-PCXR8, SKC Inc., PA, USA) was used for sampling air.

Reagents and solutions

Nitric acid (double distilled, Seastar Co., Canada) and hydrochloric acid (trace metal grade, Fisher Scientific Co., Canada) were used to prepare 0.2% HNO_3 (v/v) and 0.2% HCl (v/v), respectively, with pure water (Barnstead NANOpure system). A stock standard solution of 1000 mg/l for As(V) (H_3AsO_4 in 0.5 M HNO_3) was purchased from Merck (Germany); As(III) was prepared by dissolving 0.1320 g As_2O_3 (99.995%, Aldrich, USA) in 10 ml of 4.0% (w/v) NaOH, acidifying with 2.0 ml of conc. HCl and then diluting to 100 ml with pure water. Working standard solutions were diluted from the corresponding stock solution with 0.2% HNO_3 (v/v) for As(V) or with 0.2% HCl (v/v) for As(III). In simulating the generation of As(III) vapor in air, the stock solution of As_2O_3 was diluted in methanol instead of pure water. Quality-control solutions for As(III) were prepared from As_2O_3 (99.999% purity, Strem Chemical Co., MA, USA); while As(V) (containing 25.0 mg/l in 10% HNO_3) was purchased from High Purity Standards (HPS#290608, SC, USA). The SRM 1643d synthetic water standard (NIST, USA) contained 56.02 ± 0.73 $\mu\text{g/l}$ of As(V) in 0.5 M HNO_3 . A DMPS stock aqueous solution (300 mg/l) was prepared from 2,3-dimercaptopropane-1-sulfonate (95%, Aldrich Co., USA) weekly. Ammonium acetate (GR, Merck) and acetic acid (99.99%, Aldrich) were used to prepare an acetate buffer (2.0 M of pH 5.5). An aqueous solution containing 25 g/l of Ni^{2+} was prepared from $\text{Ni}(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$ (99.999%, Aldrich). Other chemicals were of analytical reagent grade.

Samples of drinking water

Groundwater was collected from Tunghai University without any treatment; it was then treated with softening and distillation; pure water was used as test samples. Each sample was acidified to pH < 2 with conc. HCl for testing As(III) and with conc. HNO_3 for testing As(V). One sample of Tunghai groundwater (pH about 6.0) was unacidified for a comparative study.

Analytical procedure for As(III) in water

Since the optimum pH of this study was around 5.5,⁵ we had to pre-adjust the pHs of those acidified samples to about 5.0 – 5.5 by $\text{NH}_3(\text{aq})$. After the addition of the DMPS solution (2 ml, 0.60 mg) and ammonium acetate buffer (2.0 mmol, pH 5.5) to unknown samples (20 ml), or appropriate amounts (0 – 200 ng) of As(III) standard solutions, the mixture was adjusted to 25.0 ml with pure water. The solution was mixed and allowed to stand at room temperature for about 1 h. The As(III)-

Table 1 Suitable temperature program for As(III) in methanol solution by GFAAS

Step	Temperature/°C	Duration/s
Drying	80 – 120	30
Ashing	150 – 700	30
	1200 – 1200	15
Atomization ^a	2800 – 2800	3
Cleaning	2800 – 2800	5

a. Flow was stopped during atomization step.

DMPS complex was concentrated on two conditioned Sep-Pak C_{18} cartridges (#51910, Waters Associates Co., MA, USA) in series by gravitational force at a flow rate of 0.6 ml/min. The complex in each cartridge was eluted with methanol and adjusted to 2.00 ml. After the addition of a 125 μl of Ni^{2+} solution (2.0 mg Ni^{2+}), a portion (20 μl) of the mixture was introduced into a graphite cuvette by a microsyringe (50 μl , Hamilton Co., USA) and atomized according to a temperature program (Table 1). The peak heights in the absorbance were used for a quantitative analysis.

Analytical procedure for As(V) in water

As(V) (up to 100 ng) in 0.2% HNO_3 (v/v) was pre-reduced to As(III) by adding 0.5 ml of L-cysteine (1% (w/v) in 0.2% HNO_3 (v/v)) and allowing the mixture to stand at room temperature for about 1 h. After the pH of the solution was adjusted to 5.0 – 5.5 with $\text{NH}_3(\text{aq})$, the analytical procedure for As(III) was again performed, except that five C_{18} cartridges in series were required. The total amount of [As(III)+As(V)] was determined. The amount of As(V) was obtained by subtracting As(III) from [As(III)+As(V)].

Simulation for generating As(III) vapor in air

Twenty-five milliliters of absorbing solution containing 0.60 mg DMPS and 2.0 mmol ammonium acetate buffer (pH 5.5) were pipetted to each of two 30-ml impingers (#1 and #2) connected in series, as shown in our previous paper.²² A known amount (100 ng) of As(III) (As_2O_3 in 100 μl methanol solution) was introduced into a small glass chamber (20 ml). An IR lamp (200 W, Philips Co.) was set above the chamber and turned on. During the first 10 min, the methanol vaporized completely and was carried to the impingers by a very small air pressure (1 – 2 air bubbles/10 s). Then, the air flow rate was increased to 250 ml/min and maintained for at least 2 h to carry the vaporized As(III) vapor into the absorbing solutions. The As(III)-DMPS complex in each impinger was analyzed using the procedure described above. The residue of As(III) in the chamber was extracted by adding 15 ml of fresh absorbing solution and mixing for about 1 h; then, the amount was measured as mentioned above.

Table 2 Selective retention for As(III)-DMPS complex on C₁₈ cartridges

Amount of arsenic prepared/ng		Amount of arsenic found/ng								
		Without L-cysteine			With L-cysteine					
As(III)	As(V)	#1	#2	Total ^a	#1	#2	#3	#4	#5	Total ^a
100	0	95.3	6.0	101.3	43.9	34.4	20.9	2.7	0	101.9
50	50	50.4	1.0	51.4	47.2	34.2	17.6	2.2	0	101.2
0	100	0	0	0	36.6	26.1	23.3	11.6	3.7	101.3

a. The mean of triplicates.

Results and Discussion

Selective retention for As(III)-DMPS complex on C₁₈ cartridges

Three pairs of 25 ml aqueous solutions were prepared containing 100 ng of As(III), 50 ng of As(III)+50 ng of As(V), and 100 ng of As(V). In each pair, one was left untreated and one was treated with L-cysteine, respectively. Table 2 shows that when the solutions were untreated with L-cysteine, only the amounts of As(III) (in the As(III)-DMPS complex) could be retained on two C₁₈ cartridges, while As(V) could not. When the solutions were treated with L-cysteine (H₂NC(H)(CH₂SH) COOH), some portion of As(III) might have been complexed with L-cysteine through -SH bonding besides DMPS. The complex of As(III)-cysteine might have been slightly more polar than As(III)-DMPS, and four C₁₈ cartridges were required to retain these two complexes. As(V) was pre-reduced to As(III) in the presence of L-cysteine while the L-cysteine itself might be oxidized to L-cystine (H₂NC(H)(COOH) CH₂S-SCH₂C(H)(COOH)NH₂).^{20,23} According to the distribution of the elution profile in Table 2, the polarity of the As(III)-cystine complex (through -S-S- bonding) might have been larger than that of As(III)-cysteine and As(III)-DMPS. Hence, five C₁₈ cartridges were required to retain the complexes of As(III)-DMPS+As(III)-cystine, and the total amounts of As(III)+As(V) could be quantitatively determined after treating with L-cysteine. In another experiment, an organic arsenic complex (dimethylarsenic acid-DMPS) could not be retained on C₁₈ cartridges either. The efficiency of the concentration procedure for 50 – 200 ng of As(III) as As(III)-DMPS complex on two Sep-Pak C₁₈ cartridges was in the range of 98.0 – 99.5%, in which about 96% of As(III) was retained in the first cartridge.

Temperature program

The effect of the final ashing temperatures (1000 – 1800°C for 10 – 30 s) and atomization temperatures (2400 – 2800°C for 1 – 5 s) on the absorbance was successively tested with 0.96 ng As(III) in 20 µl concentrated methanol solution. The results indicated that a final ashing temperature of 1200°C for 15 s and an

Fig. 1 Effect of DMPS on the recovery of 100 ng As(III): (a) through the concentration procedure from 25-ml aqueous solution; (b) directly prepared in methanol (2.00 ml).

atomization temperature of 2800°C for 3 s are suitable, as tabulated in Table 1.

Amount of DMPS, acetate buffer and Ni²⁺ used

Two sets of experiments were performed in order to study the stoichiometric molar ratio of DMPS versus As(III). The first one was prepared in 25 ml aqueous solutions containing 100 ng of As(III) (*i.e.* 1.3 nmol), 2.0 mmol ammonium acetate buffer (pH 5.5), and various amounts (0.0 – 1.50 mg, *i.e.*, 0 – 6.6 µmol) of DMPS, and then going through the concentration procedure. Figure 1(a) indicates that about 100 ng (100% recovery) for As(III) was found when the molar ratio of DMPS versus As(III) was in the range of 2000 – 5000 (*i.e.*, 0.60 – 1.5 mg of DMPS). Hence, 0.60 mg of DMPS was selected. The second experiment was pipetting 0.16 mmol ammonium acetate buffer (a proportional amount) and various molar ratios (0 – 2000) of DMPS versus As(III) (1.3 nmol) into small test tubes (5.00 ml); this was then made to 2.00 ml with methanol. After mixing and standing for about 1 h, 2.0 mg of Ni²⁺ was added and a portion (20 µl) was atomized. Figure 1(b) indicates that about 100% recovery of As(III) was obtained when the molar ratio of DMPS versus As(III) was in the range of 5 – 2000. The experimental results suggest the ratio [As(III)]:[DMPS]=1:5.

The amount (0 – 10.0 mmol) of ammonium acetate buffer at pH 5.5 was varied; the results indicated that all amounts in the range 2.0 – 10.0 mmol had the same

large absorbances. Hence, 2.0 mmol was selected. The effect of Ni^{2+} (0 – 50 μg) on the absorbance of 0.96 ng As(III) in a 20 μl injection showed that all amounts in the range 20 – 50 μg had the same large absorbances. Hence, 20 μg of Ni^{2+} in 20 μl was selected.

Reduction condition used for As(V)

The effect of the amount of L-cysteine (0.10 – 1.0 ml of 1% (w/v) in 0.2% HNO_3 (v/v)) and the reaction time (0.5 – 1.5 h) on the reaction of As(V) (100 ng) to As(III) was studied. The results indicated that when the amount of L-cysteine was not sufficient (<0.5 ml of 1% (w/v)) or the reaction time was not long enough (<1 h), the reduction of As(V) to As(III) was incomplete. Since five C_{18} cartridges in series were required to obtain satisfactory recoveries for 100 ng of As(V), we did not attempt to reduce larger amounts than this.

Calibration graphs, method detection limits (MDL), and accuracy tests for As(III) and As(V) in water

Following the proposed method, a typical calibration graph for As(III) was $y=0.0234x-0.0017$ with a linear range of 12.5 – 200 ng As(III) and a correlation coefficient of 0.9995. The MDL for As(III) and As(V) were determined as the concentration corresponding to three-times the standard deviation of twelve replicates using 20 μl of a mixture containing Ni^{2+} and methanol obtained from the first C_{18} cartridge concentrated from a 25 ml aqueous solution containing 0.5 $\mu\text{g/l}$ of As(III) and As(V), respectively. The MDL (3σ) from the mean of six determinations were found to be 0.11 ± 0.01 $\mu\text{g/l}$ and 0.15 ± 0.02 $\mu\text{g/l}$ for As(III) and As(V), respectively, for a 25-ml aqueous solution, or an amount of 2.8 ng As(III) and 3.8 ng As(V). Although the MDL values of As(III) and As(V) obtained in this work were higher than those (0.001 $\mu\text{g/l}^{11,12}$, 0.011 $\mu\text{g/l}^{13}$ and 0.012 $\mu\text{g/l}^{14}$ for As(III); 0.010 $\mu\text{g/l}^{11}$, 0.018 $\mu\text{g/l}^{13}$ and 0.062 $\mu\text{g/l}^{14}$ for As(V)) reported elsewhere, those values might still be useful in practice for the maximum allowable contamination of 10 $\mu\text{g/l}$ in drinking water.

The accuracies of the proposed method were checked by testing a QC sample (As_2O_3 , Strem Chemical Inc.) for As(III); a QC sample (wastewater, HPS#290608) and the certified synthetic water (SRM 1643d) for As(V). The results of 97.6 – 99.3% were obtained, as listed in Table 3. The %RSD at 100 ng of As(III) and As(V) were within 4.0% and 2.5%, respectively ($n=3$).

Effect of foreign ions on As(III) recovery

For a 25 ml aqueous solution containing 4.00 $\mu\text{g/l}$ of As(III), the effects of Ca^{2+} , Mg^{2+} , Fe^{2+} , Mn^{2+} , Na^+ , and Cl^- on the absorbance of arsenic could be tolerable up to 50, 400, 0.50, 200, 900, and 1400 mg/l with 99, 98, 97, 98, 95, and 95% recoveries, respectively.

Determination of As(III) and As(V) in drinking water

The concentrations of As(III) and As(V) in Tunghai groundwater, softened-distilled water, and pure water were determined by the proposed method and were

Table 3 Accuracy tests for As(III) and As(V) in drinking water

Species	Amount of As/ng		Accuracy ^d , %
	Prepared	Found ^d	
As(III)	50.0 ^a	48.8 $\pm$ 1.2	97.6 $\pm$ 2.5
	100 ^a	98.9 $\pm$ 2.8	98.9 $\pm$ 2.8
	200 ^a	195.8 $\pm$ 7.8	97.9 $\pm$ 4.0
As(V)	100 ^b	98.4 $\pm$ 2.5	98.4 $\pm$ 2.5
	50.5 $\pm$ 0.7 ^c	50.2 $\pm$ 0.4	99.3 $\pm$ 0.8
	100.8 $\pm$ 1.3 ^c	99.2 $\pm$ 0.5	98.5 $\pm$ 0.5

a. Prepared from a QC sample (As_2O_3 , Strem Chemical Inc.).

b, c. Prepared from a QC sample (wastewater, HPS#290608) and the standard reference material (SRM 1643d, synthetic water), respectively.

d. Mean of three determinations with standard deviation.

found as N.D. (<0.11 $\mu\text{g/l}$ for As(III) and <0.15 $\mu\text{g/l}$ for As(V)) both in acidified samples and an unacidified Tunghai groundwater (20 ml). After spiking As(III) (50.0 or 100 ng) and As(V) (100 ng) to samples, respectively, the average recoveries from triplicates were 98.5 – 102% for the unacidified sample (with %RSD within 2.8%) and 95.7 – 101% for the acidified samples (with %RSD within 4.3%).

Feasibility of collecting As(III) vapor by use of an absorbing solution

The absorbing efficiencies in impinger #1 for As(III) vapor simulated from 100 ng of As(III) in methanol ranged from 34.2 to 60.8% when the IR irradiation time was varied from 0.5 to 4 h at an air flow rate of 250 ml/min (except for one at 350 ml/min) as listed in Table 4. The results indicated that an IR irradiation time of at least 2 h was required to vaporize most of the As_2O_3 and to obtain a 60% collection efficiency in impinger #1; however, the residue portion found from the glass chamber was 11 – 13% and the amount found in impinger #2 was N.D. (<2.8 ng As(III)). The total recovery of As(III) by this generating technique was only about 72%. This might be because As_2O_3 has a high melting point (275°C) and a high sublimation point (313°C);⁷ therefore some portion of the As_2O_3 still remained on the glass chamber and was not evaporated under the IR irradiation (about 130°C) nor extracted by a fresh absorbing solution (15 ml). The % RSD for collecting As(III) vapor in impinger #1 with an IR irradiation for 2 h was 4.5% from a triplicate. From the result of absorbing in impinger #1, the collection of As(III) vapor in air might be feasible.

Determination of As(III) contents from air samples

Three pairs of air samples were collected from inside our laboratory (120 l), outside our laboratory (120 l), and on the farm of Tunghai University (120 l), which is about 400 m away from the incinerator of Taichung Veteran's Hospital, respectively, together with a corresponding blank. The obtained results were N.D.

Table 4 Collection efficiency for simulated As(III) vapor

Introduced	Amount of As(III)/ng			Total recovery, %	IR irradiation time ^b /h	Volume of air/l
	Imp.#1	Imp.#2	Chamber			
100	34.2±1.6	N.D. ^c	35.2±1.1	69.4	0.5	30 ^d
100	44.2±2.2	N.D.	26.7±1.8	70.9	1	30 ^d
100	60.3±2.7	N.D.	11.0±2.6	71.3	2	30 ^d
100	60.5±2.0	N.D.	11.7±2.3	72.2	2	42 ^f
100	60.8±1.8	N.D.	11.2±2.5	72.0	4	60 ^e

a. Mean of triplicates with standard deviation. b. IR lamp further irradiation time after 10 min. c. N.D. <2.8 ng. d, e and f. Air flow rate at 250 ml/min for 2 h; for 4 h; and 350 ml/min for 2 h, respectively.

(which is <2.8 ng of As(III) for an air volume of 120 l, or about <23 ng/m³).

The authors thank the National Science Council of the Republic of China for financial support (NSC 86-2113-M-029-007).

References

- W. P. Tseng, H. M. Chu, S. W. How, J. M. Fong, C. S. Lin and S. Yeh, *J. Nat. Cancer Inst.*, **40**, 453 (1968).
- C. J. Chen, Y. Chuang, T. M. Lin and H. T. Wu, *Cancer Res.*, **45**, 5895 (1985).
- A. H. Smith, H. M. Duggan, R. Wood, M. J. Kosnett and M. T. Smith, *Environ. Health Perspect.*, **97**, 259 (1992).
- S. L. Chen, S. R. Dzung, M. H. Yang, K. H. Chiu, G. M. Shieh and C. M. Wai, *Environ. Sci. Technol.*, **28**, 877 (1994).
- Y. C. Chen, H. C. Wang, S. Y. Cheng, C. H. Yen and M. S. Kuo, *J. Environ. Prot. Soc. [R.O.C.]*, **19**, 14 (1996).
- F. W. Pontius, *J. AWWA*, **Feb.**, 48 (1995).
- NIOSH method 7901, "Arsenic Trioxide, as As", 1994.
- M. Sperling, X. Yin and B. Welz, *Spectrochim. Acta*, **46B**, 1789 (1991).
- B. Welz and M. Sucmanova, *Analyst* [London], **118**, 1417 (1993).
- T. R. Rude and H. Puchelt, *Fresenius' J. Anal. Chem.*, **350**, 44 (1994).
- M. Hasegawa, Y. Sohrin, M. Matsui, M. Hojo and M. Kawashima, *Anal. Chem.*, **66**, 3247 (1994).
- W. M. Mok and C. M. Wai, *Anal. Chem.*, **59**, 233 (1987).
- C. J. Hwang and S. J. Jiang, *Anal. Chim. Acta*, **289**, 205 (1994).
- M. L. Magnuson, N. T. Creed and C. A. Brockhoff, *J. Anal. At. Spectrom.*, **11**, 893 (1996).
- H. V. Aposhian, C. H. Tadlock and T. E. Moon, *Toxicol. Appl. Pharmacol.*, **61**, 385 (1981).
- D. F. Moore, C. A. Ocallaghan, G. Berlyne, C. S. Ogg, H. A. Davies, I. M. House and J. A. Henry, *J. Neurol. Neurosur. Psych.*, **57**, 1133 (1994).
- R. D. Ediger, *At. Absorpt. Newsl.*, **14**, 127 (1995).
- D. C. Manning and W. Slavin, *Appl. Spectrosc.*, **37**, 1 (1983).
- W. L. Jolly, "The Principles of Inorganic Chemistry", p. 199, McGraw-Hill, New York, 1976.
- H. Chen, I. D. Brindle and X. C. Le, *Anal. Chem.*, **64**, 667 (1992).
- Y. L. Feng and J. P. Cao, *Anal. Chim. Acta*, **293**, 211 (1994).
- H. C. Wang, Y. C. Hwang, C. J. Hsieh and M. S. Kuo, *Anal. Sci.*, **14**, 983 (1998).
- L. Stryer, "Biochemistry", 2nd ed., pp. 15 - 18, W. H. Freeman and Company, San Francisco, 1981.

(Received November 9, 1998)

(Accepted April 22, 1999)