specific conformation for prion protein. Conversion from the cellular form (PrP^C) to the scrapie form (PrP^{Sc}) is essentially a misfolding process without changing in covalent structure. Thus it may be possible to intervene in this process by introducing a small compound which can act as a chemical chaperon capable of stabilizing the PrP^C conformation [1]. Initially we selected the more than 200 compounds from *in silico* screening based on the PrP^C structure and finally discovered 24 efficient anti-prion compounds with diverse chemical structures from *ex vivo* experiment [2]. Surface plasmon resonance studies reveals that the binding affinities of compounds roughly correlated with their anti-prion activities. This result indicates that identification of compounds that acts as chemical chaperons is one of the efficient strategy for anti-prion drug discovery. Existence of the exception indicating the involvement of some modulation factors other than the direct binding with PrP^C. The categorization of these diverse compounds would facilitate the anti-prion drug discovery and the elucidation of the pathogenic conversion mechanism. [1] K. Kuwata et al. (2007) PNAS 104, 11921. [2] Hosokawa-Muto, Kamatari, Nakamura, Kuwata (in preparation).

3P-049 QM/MM 法によるアデニル酸キナーゼの反応機構に関する研究
Theoretical study on the reaction mechanism of adenylate kinase by QM/MM method

Kenshu Kamiya, (Dept Phys, School of Sciences, Kitasato Univ)

Adenylate kinase catalyzes the reaction: $ATP + AMP + Mg^{2+} \rightarrow ADP + ADP + Mg^{2+}$. In order to elucidate the nature of the reaction, we have investigated the reaction mechanism of the system using molecular mechanics, molecular dynamics and quantum chemical methods. We constructed the complex structure of enzyme and substrates, ATP and AMP with Mg ion based on the PDB data 2AKY. After the full optimization of the whole molecule in vacuo, we put water molecules around catalytic center Mg ion within radius of 12 Å with spherical boundary condition centered on Mg ion, and we put harmonic constraint potentials on the surrounding atoms in the boundary region. Some model structures were sampled from 900K MD simulation followed by structural energy minimizations. In order to obtain the reasonable model structure of the transition state of the reaction, we performed molecular mutation calculation, in which values of molecular mechanics parameters of the reactant gradually turned into those of the product during MD simulation. As the result we were able to obtain the reasonable model structure of the transition state of the reaction. We truncated surrounding atoms farther than about 10 Å from the reaction center, then we optimized the transition structure with ONIOM (QM/MM) method as well as the reactant and the product states. The reaction path of SN2 type inversion of phosphate group were retained by hydrogen bondings with surrounding atoms.

3P-050 単粒子解析における最尤法に基づく粒子拾い上げ
Maximum likelihood approach for picking-up process in single particle analysis
M. Kawaka, Y. Maruyama and C. Sato, (National Institute of Advanced Industrial Science and Technology (AIST))

We present a new method for generating a set of high quality particles in single particle analysis (SPA) from electron microscope images. The method uses maximum-likelihood (ML) approach which determines positions of particles by applying ML estimations using reference particle data. ML approach has some advantages over the conventional methods that have been used in the picking-up process in SPA; ML approach can minimize statistical errors in the determination of particle positions and excise the most reliable particle data set under some assumed conditions. However, there are some difficulties to implement the ML calculations; 1) Gaussian distributions of some parameters are required for the fast calculations, 2) complete survey of the parameter space is almost impossible due to the limited computational resources. This method was examined for its picking-up efficiency, i.e. accuracy in recognizing particles and noise tolerance, using model image sets with various signal-to-noise ratios and with electron microscope images of the negatively stained Transient receptor potential melastatin type 2. The method yielded high performance in particle recognitions and produced a set of particles accurate enough for processing data in the following steps in SPA. These results showed that ML-based picking-up method is promising even for image data with extremely low SNR or contrast images in SPA.

3P-051 3重共鳴 4次元 NMR への非線形サンプリングの応用
Applications of nonlinear sampling scheme to four dimensional triple resonance NMR spectroscopy.

Yoshiki Shigemitsu(1,2), Yuusuke Tsuchie(1), Masaki Mishima(1,2), Daniel Nietlispach(3), Markus Waechli(4) and Yutaka Ito(1,2). (1: Department of Chemistry, Tokyo Metropolitan University, 2: CREST/JST, 3: Department of Biochemistry, University of Cambridge, 4: Bruker Biospin)

Despite its potential advantages in analysis, long duration of measurement and limited digital resolution in indirectly observed dimensions prevent 4D NMR experiments from being used in protein NMR projects routinely. Applications of recently proposed new acquisition schemes for speeding up multidimensional NMR experiments are therefore expected to be extremely beneficial for 4D experiments. Nonlinear sampling for indirectly acquired dimensions in combination with maximum entropy (MaxEnt) reconstruction has been shown to provide significant time savings in the measurement of multidimensional NMR experiments. In this presentation, we demonstrate the benefits of this nonlinear sampling and 3D MaxEnt processing in 4D triple-resonance experiments. 4D HCC(CO)NH and 4D ¹³C/¹⁵N separated NOESY data were acquired with 512 (¹H) x 32 (¹H) x 24 (¹³C) x 8 (¹⁵N) linearly sampled complex points on T.

thermophilus HB8 TTHA1718 protein (66a.a.). For both 4D experiments we succeeded in reconstructing spectra of equivalent quality to the reference spectra from eminently reduced sampling points. The time gained from this scheme can be used to increase the number of indirect data points. Resolution enhancement of 4D spectra was consequently achieved by measuring data with extended acquisition times for indirect dimensions with nonlinear sampling scheme. We also evaluated the expected artefacts in 4D spectra arising from nonlinear sampling and MaxEnt processing, such as the mis-calibration of intensities and the emergence or disappearance of cross peaks.

3P-052 タンパク質の主鎖トポロジーとフォールディングキネティクスとの関係—3次元格子タンパク質の円順列変異体に統計力学モデルを適用して—

Relationship between folding topology and kinetics using a statistical mechanical model for 3D lattice proteins to circular permutants

Haruo Abe (1), Hiroshi Wako (2). (1: Dept. of Electrical Engineering, Nishinippon Inst. of Tech. 2: School of Social Sciences, Waseda Univ.)

In order to understand protein folding mechanisms, we analyzed the relationship between the contact order (CO) and folding rates (k_f) in circular permutants of 48-mer 3D lattice proteins. At first, we examined how the interaction energies between amino acid residues in 48 circular permutants are distributed with respect to their distances along the chain. The distances between the amino acid residues along the chain, i.e., the distributions of short-, medium-, and long-range interactions are a key point of CO. Then, we calculated k_f using a statistical mechanical model, which we referred to as A-W model^{1,2,3)}. Finally, we analyzed the correlation between CO and ln k_f. We have grouped the 48 circular permutants into several types according to their core structures to analyze the relationship between CO and ln k_f. We observed an excellent correlation between CO and ln k_f in each type of circular permutants and found that, although the contact order contributes to folding rate essentially, the interaction energies between amino acid residues distant along the chain also play an important role. We will discuss them in detail in the meeting. 1)H.Abe & H.Wako, J. Phys. Soc. Jpn. 73(2004)1143. 2)H.Abe & H.Wako, Phys. Rev. E 74(2006) 011913. 3) H.Wako & H.Abe, J. Phys. Soc. Jpn. 76(2007)104801.

3P-053 単粒子解析で分子構造の違う画像を分離する方法
Separating images in different conformations of molecules on single particle analysis

Yutaka Ueno(1), Patrick Schultz(2), Gavor Papai(2). (1: Neuroscience, AIST; 2: Departement de Biologie et Genomique Structurales, IGBMC)

Using electron microscopy projection images of biological molecules, a computational method was developed to separate different conformations of the same molecule or different particles. Purified protein molecules may exhibit different conformations in solution which affects the resolution of the 3D reconstruction in which all views are merged. Resolution can only improve if the different conformations are sorted out. Although fundamental limitations exist on heterogeneity determination when only projection images are analyzed, a statistical method to build reasonable 3D models by sorting images into different models would be helpful. Furthermore, pioneering studies on ribosome complexes showed that different conformational states can be extracted from a heterogeneous data set and contribute to explain its molecular mechanism. After particle images are classified into classes, a score is introduced to measure how well two given class averages fit to the same 3-D model. In our method, the similarity between common lines of the projections, known as the cross sinogram correlation was used. Particle images of the classes are placed on a virtual map where similar images are connected to form a graph. Then, a modern algorithm for the graph-cut problem is applicable to separate images. Using a test data set containing mixed projection images of simple phantoms, a protocol to obtain correct separation was investigated. The method was also applied to an image data set of RNA polymerase whose conformational flexibility was reported previously.

3P-054 マイクロビーズとマイクロ流路を組み合わせた新規プロテインアレイの開発

A novel protein array using microbeads aligned in a microfluidic chip
Yoshihiro Ooe (1), Yasuhiro Sasuga (2), Osamu Ohara (3), Yoshie Harada (2,4). (1: Dep Medical Genome, Graduate School of Frontier Science, The Univ. of Tokyo; 2: The Tokyo Metropolitan Inst. of Med.Sci.; 3: Kazusa DNA Inst.; 4: iCeMS, Kyoto Univ.)

Protein array is a powerful means to investigate protein-protein interactions. Yet current protein arrays are not versatile due to their low sensitivity (1ng/ml) and cost-ineffectiveness.

In this study, we have developed a sensitive and cost-effective protein array using a commercial fluorescence microscope. The protein array has aligned antibody-immobilized microbeads (5µm in diameter) inside a polydimethylsiloxane (PDMS) microfluidic chip. The minimum concentration required for fluorescence detection was determined to be several tenths of pM (about 1pg/ml) using fluorescently-labeled glutathione-S-transferase (GST) to the protein array having α-GST antibody immobilized microbeads.

Firstly, we tried detecting a recombinant protein expressed in cultured cells. We extracted cytoplasmic components of PC12 cell expressing green fluorescent protein (GFP) and labeled them with amino group reactive fluorescent dye. The labeled product was applied to the protein array having α-GFP antibody microbeads, α-β-actin antibody microbeads as positive control and α-IgE antibody microbeads as negative control. Only α-GFP and α-β-actin antibody