1P302 超音波高速 AFM の開発に向けた基礎研究 2
Pilot study 2 for the development of high-speed ultrasonic AFM

Tomofumi Saito¹, Noriyuki Kodera², Toshio Ando^{1,2} (¹*Sch. Math. & Phys., Inst. Sci.*, ²*Bio-AFM Frontier Research Center, Inst. Sci. & Eng., Kanazawa Univ.*)

We have developed high-speed AFM (HS-AFM) that enables us to directly visualize dynamic structural changes of protein molecules at high spatiotemporal resolution. So far, various biological processes have been directly filmed. However, the observations by HS-AFM are generally limited only to phenomena occurring on relatively hard surfaces, meaning that we can observe neither objects placed on a soft surface nor objects placed under a surface. Ultrasonic techniques must be useful in breaking this limitation because they have been widely used for non-invasiveness imaging of objects lying on/under a surface. Thus, we are currently attempting to find a way to combine our HS-AFM with ultrasonic techniques. In the presentation, we will report our pilot study on this subject.

1P303 Real-time observation of amyloid fibril formation of yeast prion Sup35 by high-speed atomic force microscopy

Liwen Zhu¹, Hiroki Konno¹, Momoko Okuda², Noriyuki Kodera¹, Toshio Ando¹, Hideki Taguchi² (¹*Bio-AFM Frontier research center, Kanazawa University*, ²*Department of Biomolecular Engineering, Graduate School of Biosciences and Biotechnology, Tokyo Institute of Technology*)

Yeast prion [PSI⁺] is caused by ordered protein aggregates, called amyloids, of Sup35 protein. To investigate the mechanism of how the amyloids elongate and then propagate is critical to unravel the amyloid biology. Although there have been some studies of real-time amyloid formation of Sup35, especially using fluorescence probes, the detailed mechanism of the amyloid formation still remains unclear. We used the high-speed atomic force microscope (HS-AFM) to investigate the dynamic behaviors of the Sup35 fibrils. We have developed procedures to watch real-time fibril formation of the recombinant N-terminal and medium domains of Sup35 in liquid at a high spatial resolution.

1P304 高速 AFM によるバクテリアの高分解能観察
Nanoscale investigation on bacterial cell surface using high-speed AFM

Hiroki Watanabe¹, Carriou David¹, Takayuki Uchihashi^{1,2}, Toshio Ando^{1,2} (¹*Dep. Phys., Col. of Sci. and Engr., Kanazawa Univ.*, ²*Bio-AFM Frontier Res. Center*)

A bacterial cell has variety of proteins and structures on the surface, such as membrane proteins, peptidoglycan layers, a flagellum and secretion apparatus. Cryo-electron microscopy can visualize their static structures and also optical microscopy can observe localization of labeled target molecules. However, there have been no techniques which can directly observe the structure of cell surface. Recently it was demonstrated that high-speed atomic force microscopy (HS-AFM) can directly visualize dynamics of protein molecules on a bacterial cell surface. Here, we applied HS-AFM to visualize the dynamics and structures on gram-negative bacteria. We will show high-resolution images of the outer membrane covered by a net structure and a lysis process induced by a detergent.

1P305 高速 AFM / 光学顕微鏡複合機
Combined system of High-speed-AFM and optical microscopy

Shingo Fukuda¹, Takayuki Uchihashi^{1,2}, Ryota Iino³, Toshio Ando^{1,2} (¹*Department of Mathematics and Physics, Grad School of Natural Science and Technology, Kanazawa University*, ²*Bio-AFM Frontier Reserch Center, College of Science and Engineering, Kanazawa University*, ³*Department of Applied Chemistry Grad School of Engineering The University of Tokyo*)

Combined system of high-speed atomic force microscopy (HS-AFM) and total internal reflection fluorescence microscopy (TIRFM) has been developed. This novel instrument can provide a way to correlate conformational change of protein to ligand-binding events, specify the tip position on a large object such as a living cell, and identify target molecules even in a complex system. Also unlike conventional combination of AFM and optical microscopy, our system can simultaneously capture AFM and TIRFM images at single molecular level. The capability of this combined microscope is demonstrated by simultaneous HS-AFM/TIRFM imaging of myosin V walking on an actin filament.

1P306 高速 AFM による ClpB の構造ダイナミクスの観察
Conformations and dynamics of ClpB hexameric ring observed by high-speed AFM

Takayuki Uchihashi^{1,2}, Yo-hei Watanabe³, Ryota Iino⁴, Hiroki Watanabe¹, Takashi Yamasaki³, Toshio Ando^{1,2} (¹*Dept. Phys., Kanazawa Univ.*, ²*Bio-AFM Frontier Research Center, Kanazawa Univ.*, ³*Dept. Biol., Konan Univ.*, ⁴*Dept. Appl. Chem., Univ. of Tokyo*)

The molecular chaperon ClpB is a ring-forming AAA⁺ machine that rescues aggregated proteins. So far, conformations of ClpB in a hexameric ring depending on nucleotides have been investigated by electron microscopic single-particle analysis and biochemical analysis. Here we applied high-speed atomic force microscopy (HS-AFM) to directly visualize conformations and dynamics of the ClpB ring. The HS-AFM images revealed that the ring shows deformed, 2-fold rotational symmetry without nucleotide, and one of the interfaces between subunits at axis of symmetry repeats reversible opening and closing. On the other hand, the ring showed 6-fold rotational symmetry in high ATP. These dynamic conformational changes in the ClpB ring would play roles for protein disaggregation.

1P307 動的な DNA コンピューティングを実現するための AND ゲートモジュールの開発
Development of AND gate module for dynamic DNA computing

Takakshi Nukada, Koh-ichiroh Shohda, Akira Suyama (*Grad. Sch. Arts and Sciences, Univ. Tokyo*)

RTRACS (Reverse-transcription and TRanscription-based Autonomous Computing System) is a modular biomolecular computing system composed of DNA, RNA and enzymes, whose mechanism is based on retroviral replication. The modules of RTRACS communicate with each other through RNA for the sophisticated computational operations. Here, we report AND gate module for RTRACS that produces output RNA after receiving input RNA and changes output RNA dynamically depending on input RNA. The module allows RTRACS to perform a time-dependent computation. It is expected that other RTRACS modules dynamically worked are also capable of being developed in the same way of the AND gate module.