ION-TRANSFER VOLTAMMETRY WITH THE INTERFACES BETWEEN
POLYMER-ELECTROLYTE GEL AND ELECTROLYTE SOLUTIONSToshiyuki OSAKAI, Tadaaki KAKUTANI, and Mitsugi SENDA[®]Department of Agricultural Chemistry, Faculty of Agriculture,
Kyoto University, Kyoto 606

The transfer of tetramethylammonium and picrate ions at the poly(vinyl chloride)-nitrobenzene gel/water and agar-water gel/nitrobenzene interfaces was studied by means of cyclic voltammetry. Well-defined cyclic voltammograms of ion transfer were obtained for both ions with these polymer gel electrodes. The results indicate that the polymer gel electrode is promising for use in electrochemical analysis of ions.

Recently a growing number of papers have appeared on electrochemical studies using polarographic and voltammetric techniques on ion transfer at the interface between two immiscible electrolyte solutions, such as the water/nitrobenzene interface (for reviews, see Refs. 1 to 4). Analytical application of these studies using a liquid/liquid interface as an electrode interface has been described¹⁾⁻⁶⁾. From a practical point of view, however, liquid/liquid interfaces or liquid electrodes are not necessarily easy to handle. In this study we have tried to solidify the liquid electrode (one of the liquid phases) with polymer materials, such as poly(vinyl chloride) for nitrobenzene and agar for water, as a matrix. The resulting poly(vinyl chloride)-nitrobenzene (PVC-NB) gel/water and agar-water gel/nitrobenzene interfaces work quite satisfactorily to give well-defined voltammograms of the transfer of tetramethylammonium (TMA^+) and picrate (Pic^-) ions. Certain polymer or polymer-gel effects on the ion-transfer process have been suggested. The results indicate that the polymer gel electrode is promising for use in the electrochemical analysis of ions.

EXPERIMENTAL

Chemicals

Lithium picrate (LiPic) was prepared according to the reported method⁷⁾. Tetrabutylammonium picrate (TBAPic) was prepared by the equimolar addition of an aqueous solution of tetrabutylammonium bromide to an aqueous solution of sodium picrate, followed by washing with distilled water; the resulting crude crystal was recrystallized twice from ethanol with water and dried at 50°C in a vacuum. Analytical grade nitrobenzene (Wako Pure Chemical Industries, Ltd.) was shaken with neutral alumina overnight and filtered; the filtrate was used for preparing nitrobenzene solutions. Triply distilled water was used for preparing aqueous

solutions. The other chemicals used have been described elsewhere⁸⁾⁹⁾.

Preparation of the polymer gel electrodes

A PVC gel electrode: two grams of PVC powder ($\overline{DP} = 1100$, Nakarai Chemicals Co.) was dissolved in 10 cm³ of 0.1 mol dm⁻³ tetrabutylammonium tetraphenylborate (TBATPB) nitrobenzene solution at ca. 100°C; the resulting solution was put on one end of a silicone tube (0.294 cm inner diameter) and cooled at room temperature to make a PVC-NB gel electrode, as shown in Fig. 1a. An Ag/AgCl electrode was immersed in the internal solution of 0.1 mol dm⁻³ tetrabutylammonium chloride (TBACl) aqueous solution saturated with nitrobenzene. The interface between the 0.1 mol dm⁻³ TBACl aqueous solution and the PVC gel containing 0.1 mol dm⁻³ TBATPB nitrobenzene solution is nonpolarizable, the potential difference across the interface being determined by the ratio of activities of TBA⁺ ion in the two phases similarly to that across a water/nitrobenzene interface³⁾⁸⁾. The PVC-NB gel electrode was stored in a 0.1 mol dm⁻³ LiCl aqueous solution saturated with nitrobenzene.

An agar gel electrode: three grams of Difco Bacto agar was dissolved in 100 cm³ of 0.1 mol dm⁻³ LiCl aqueous solution at ca. 100°C; the resulting solution was put on one end of a glass tube (0.33 cm i.d.) and cooled at room temperature to make an agar-water gel electrode, as shown in Fig. 1b. An Ag/AgCl electrode was immersed in the internal solution of 0.1 mol dm⁻³ LiCl aqueous solution. The agar-water gel electrode was stored in a 0.1 mol dm⁻³ LiCl aqueous solution.

Fig. 1 (a) PVC-NB gel electrode and (b) agar-water gel electrode

Electrochemical measurement

In recording the voltammograms, a two-electrode cell system⁹⁾ was used, the ohmic drop due to the solution and gel resistance between two reference electrodes being compensated for by positive feedback. The solution and gel resistance was determined to be 12-19 k Ω for the PVC-NB gel electrodes and 2.4-3.6 k Ω for the

agar-water gel electrodes using a linear bridge⁹).

With the PVC-NB gel electrode, the transfer of TMA^+ and Pic^- ions across the interface between phases II and III in the following electrochemical cell was studied:

The left-hand half-cell Ag/AgCl/I/II represents the PVC-NB gel electrode of Fig. 1a. An Ag/AgCl electrode was used as the right-hand reference electrode. The aqueous solution of phase III was saturated with nitrobenzene.

With the agar-water gel electrode, the transfer of Pic^- ion across the interface between phases II and III in the following electrochemical cell was studied:

The right-hand half-cell Ag/AgCl/IV/III represents the agar-water gel electrode of Fig. 1b. The left-hand Ag/AgCl/ $0.1 \text{ mol dm}^{-3} \text{ TBACl}(\text{water})$ electrode was connected to phase II through sintered glass⁹. The nitrobenzene solution in phase II was saturated with water.

The potential applied to the cells, E , is defined by the potential of the right-hand reference electrode, E_{right} , referred to the potential of the left-hand reference electrode, E_{left} , i.e. $E = E_{\text{right}} - E_{\text{left}}$. The flow of positive charge from phase III (aqueous) to II (organic) is taken as the positive current. The electrolytic cells were thermostated at $25 \pm 0.1^\circ\text{C}$.

RESULTS

Voltammetry with the PVC-NB gel electrode

As shown in Fig. 2, TMA^+ and Pic^- ions in aqueous solution gave well-defined cyclic voltammograms at the PVC-NB gel electrode. In cyclic voltammogram A in Fig. 2, the anodic (positive current) peak corresponds to the transfer of TMA^+ ion from water to PVC-NB gel while the cathodic (negative current) peak corresponds to its transfer back to water. In cyclic voltammogram B in Fig. 2, the cathodic peak corresponds to the transfer of Pic^- ion from water to the PVC-NB gel while the anodic

Fig. 2 Cyclic voltammograms obtained with the PVC-NB gel electrode for $1.0 \times 10^{-3} \text{ mol dm}^{-3}$ TMA^+ ion (A) and Pic^- ion (B) in 0.1 mol dm^{-3} LiCl aqueous solution at the voltage scan rate of 10 mV s^{-1}

Dashed curves indicate the respective base currents.

peak corresponds to its transfer back to water. The anodic peak current, I_{pa} , for TMA^+ ion and the cathodic peak current, I_{pc} , for Pic^- ion, when corrected for the base currents, were proportional to the bulk concentrations of the ions in the aqueous phase, C_j^W , over the range from 0.5×10^{-3} to $2.0 \times 10^{-3} \text{ mol dm}^{-3}$, and to the square root of the voltage scan rate, v , over the range from 5 to 100 mV s^{-1} . Figures 3 and 4 demonstrate these relationships for TMA^+ -ion transfer. The diffusion coefficients of the ions in the aqueous phase were found to be $(1.0 \pm 0.1) \times 10^{-5}$ and $(5.7 \pm 0.3) \times 10^{-6} \text{ cm}^2 \text{ s}^{-1}$ for TMA^+ and Pic^- ions, respectively. The value for TMA^+ ion agreed well with the value of $1.2 \times 10^{-5} \text{ cm}^2 \text{ s}^{-1}$ obtained using

Fig. 3 Dependence of the anodic peak current for the TMA^+ -ion transfer from 0.1 mol dm^{-3} LiCl aqueous solution to ca. 20% w/v PVC- 0.1 mol dm^{-3} TBATPB NB gel, I_{pa} , on the bulk concentration of TMA^+ ion, C_{TMA}^W

The voltage scan rate v : (O) 5, (●) 10, (Δ) 20, ($\blacktriangle$) 50, ($\square$) 100 mV s^{-1} .

Fig. 4 Dependence of the anodic peak current for the TMA^+ -ion transfer from 0.1 mol dm^{-3} LiCl aqueous solution to ca. 20% w/v PVC- 0.1 mol dm^{-3} TBATPB NB gel, I_{pa} , on the square root of voltage scan rate, $v^{1/2}$

The bulk concentration of TMA^+ ion C_{TMA}^W : (O) 0.5×10^{-3} , (Δ) 1.0×10^{-3} , ($\square$) $2.0 \times 10^{-3} \text{ mol dm}^{-3}$.

a water/nitrobenzene interface⁸⁾. The difference between the anodic peak potential, E_{pa} , and the cathodic peak potential, E_{pc} , i.e. $\Delta E_p = E_{pa} - E_{pc}$ was 77 ± 9 mV for TMA^+ ion and 74 ± 7 mV for Pic^- ion, which were slightly larger than the theoretical value of 60 mV for the reversible wave. This discrepancy should partly be attributed to residual ohmic drop, though the effect of kinetic control cannot be fully excluded. The midpoint potentials, E_{mid} , defined by $E_{mid} = (E_{pa} + E_{pc})/2$, were 331 ± 1 mV and 301 ± 2 mV for TMA^+ and Pic^- ions, respectively, independently of the bulk concentration of the ions and the voltage scan rate. This independence indicates that the midpoint potentials can be considered nearly identical with the reversible half-wave potentials of the ions. The values of E_{mid} and I_{pa} (or I_{pc}) were reproducible within the errors of ± 1 mV and $\pm 2\%$, respectively, using different PVC-NB gel electrodes of the same construction. The PVC-NB gel electrode, when stored in a LiCl aqueous solution, was usable repeatedly for more than a week.

Voltammetry with the agar-water gel electrode

As shown in Fig. 5, Pic^- ion in nitrobenzene gave a well-defined cyclic voltammogram at the agar-water gel electrode. The anodic peak corresponds to the transfer of Pic^- ion from nitrobenzene to the agar-water gel while the cathodic peak corresponds to its transfer back to nitrobenzene. I_{pa} was proportional to the bulk concentration of Pic^- ion and the square root of the voltage scan rate (data not shown). The diffusion coefficient of Pic^- ion in the nitrobenzene phase was found to be $(4.9 \pm 0.8) \times 10^{-6} \text{ cm}^2 \text{ s}^{-1}$. ΔE_p was 76 ± 10 mV; the discrepancy from the theoretical value of 60 mV should partly be attributed to residual ohmic drop, but the effect of kinetic control cannot be fully excluded. E_{mid} was 328 ± 1 mV, independently of the bulk concentration of Pic^- ion and the voltage scan rate. The reproducibility of E_{mid} and I_{pa} using different agar-water gel electrodes of the same construction was ± 2 mV and $\pm 5\%$, respectively.

Fig. 5 Cyclic voltammogram obtained with the agar-water gel electrode for $2.0 \times 10^{-3} \text{ mol dm}^{-3}$ Pic^- ion in 0.1 mol dm^{-3} TBATPB nitrobenzene solution at the voltage scan rate of 10 mV s^{-1}

Dashed curve indicates the base current.

DISCUSSION

The present experimental results clearly indicate that the ion-transfer at the polymer gel/solution interface can be studied using electrochemical techniques equally as well as at the interface between two immiscible electrolyte solutions. From an analytical point of view, the polymer gel electrode has the advantage that it can be formed in any desired shape and size and is easy to handle.

The reversible half-wave potentials of the transfer of TMA^+ ion⁸⁾¹⁰⁾ and Pic^- ion¹¹⁾ at the interface between 0.1 mol dm^{-3} TBATPB nitrobenzene and 0.1 mol dm^{-3} LiCl aqueous solution were found to be 320 mV and 321 mV, respectively, with essentially the same electrochemical cell as Cell A except that the polymer gel phase is replaced by the solution phase, whereas the midpoint potentials, which could be considered identical with their reversible half-wave potentials, of the transfer of TMA^+ and Pic^- ions at the ca. 20% w/v PVC- 0.1 mol dm^{-3} TBATPB nitrobenzene gel/ 0.1 mol dm^{-3} LiCl aqueous solution interface were 331 mV and 301 mV, respectively. The midpoint potential of the Pic^- -ion transfer at the ca. 3% w/v agar- 0.1 mol dm^{-3} LiCl aqueous solution gel/ 0.1 mol dm^{-3} TBATPB nitrobenzene interface was 328 mV. The standard potential of j-ion transfer at the interface between two immiscible electrolyte solutions (α and β), $\Delta_{\beta}^{\alpha} \phi_j^{\ominus}$, is related to the standard Gibbs energy of the transfer of j ion from β to α phase, $\Delta G_{\text{tr},j}^{\ominus, \beta \rightarrow \alpha}$, by

$$\Delta_{\beta}^{\alpha} \phi_j^{\ominus} = - \Delta G_{\text{tr},j}^{\ominus, \beta \rightarrow \alpha} / z_j F, \quad (1)$$

where z_j is the ionic valence of j ion and F the Faraday constant. Then the reversible half-wave potential, $E_{1/2,j}^r$, is given by⁸⁾

$$E_{1/2,j}^r = \Delta E_{\text{ref}} + \Delta_{\beta}^{\alpha} \phi_j^{\ominus} + (RT/z_j F) \ln(\gamma_j^{\beta}/\gamma_j^{\alpha}) + (RT/z_j F) \ln(D_j^{\alpha}/D_j^{\beta})^{1/2}, \quad (2)$$

where ΔE_{ref} is a constant which depends only on the reference electrode system, R the gas constant, T the absolute temperature, and γ_j^i and D_j^i ($i = \alpha$ and β) are the activity coefficient and the diffusion coefficient of j ion, respectively, in the i phase. These equations can equally well be applied to the interface between a polymer-electrolyte gel and an electrolyte solution. Accordingly the difference between the reversible half-wave potentials and the midpoint potentials should be due to the difference in the standard potential of ion transfer between solution/solution and polymer gel/solution interfaces and to the difference in the diffusion coefficient and activity coefficient between the two phases. The difference in ΔE_{ref} 's value should also be taken into account for the most rigorous comparison. The standard Gibbs energy of transfer (partition coefficient) of a solute between gel phase and solvent phase has been discussed as a function of the composition of the matrix polymer and solvent of the gel¹²⁾¹³⁾. The kinetic effect of a polymer gel on the ion transfer across the interface seems interesting enough to be studied further. Incorporation of suitable ligand(s) or ionophore(s) into the polymer gel by dissolution in gel or covalent-bonding to polymer may further affect the apparent standard potential of ion transfer, rendering the polymer gel electrode ion selective for analytical purposes. Potentiometry as well as amperometry with polymer gel electrodes is feasible.

REFERENCES

- 1) J. Koryta, P. Vanýsek: "Advances in Electrochemistry and Electrochemical Engineering", Edited by H. Gerisher, C. W. Tobias, Vol. 12, p. 113 (1981). (John Wiley & Sons, Inc., New York).
- 2) M. Senda, T. Kakutani: *Hyomen*, 18, 535 (1980).
- 3) M. Senda, T. Kakutani, T. Osakai: *Denki Kagaku*, 49, 322 (1981).
- 4) J. Koryta: *Ion-selective Electrode Rev.*, 5, 131 (1983).
- 5) V. Mareček, Z. Samec: *Anal. Lett.*, 14, 1241 (1981).
- 6) V. Mareček, Z. Samec: *Anal. Chim. Acta*, 141, 65 (1982).
- 7) I. M. Kolthoff: *Can. J. Chem.*, 59, 1548 (1981).
- 8) T. Kakutani, T. Osakai, M. Senda: *Bull. Chem. Soc. Jpn.*, 56, 991 (1983).
- 9) T. Osakai, T. Kakutani, M. Senda: *Bull. Chem. Soc. Jpn.*, 57, 370 (1984).
- 10) T. Osakai, T. Kakutani, Y. Nishiwaki, M. Senda: *Bunseki Kagaku*, 32, E81 (1983).
- 11) T. Osakai, T. Kakutani, M. Senda: unpublished results.
- 12) K. Saitoh, N. Suzuki: *Bull. Chem. Soc. Jpn.*, 51, 116 (1978).
- 13) K. Saitoh, E. Ozaki, N. Suzuki: *J. Chromatogr.*, 239, 661 (1982).

Keyword phrases

polymer-electrolyte gel/electrolyte solution interface; polymer gel electrode; gel electrode; poly(vinyl chloride)-nitrobenzene gel; agar-water gel; ion-transfer voltammetry.

(Received March 1, 1984)