Extraction of Catecholamine from Mouse Tissues by Ion-exchange Chromatography and Its Determination by HPLC

H. DOHI*, S. SAKAI, K. KOHMOOTO and Y. SHODA

Department of Animal Breeding, University
of Tokyo, Tokyo 113

(Received March 1, 1985)

Abstract Catecholamine was extracted from various tissues of mice by Amberlite CG50 column chromatography with high recoveries (more than 90%). Epinephrine, norepinephrine and dopamine were separated by high-pressure liquid chromatography with a reversed-phase column and an electrochemical detector and simultaneously measured by the reading of peak heights on the high-pressure liquid chromatographic recorder. Correlation between the peak height and the amount of standards applied to the column was high ($r=0.99$). Inter- and intra-assay coefficients of variation were less than 2.6 and 3.0%, respectively. The detection limit (signal/noise >2) was 75 fmol for norepinephrine and 100 fmol for epinephrine and dopamine. When tissue samples were applied onto alumina columns, some unidentified substances were extracted with catecholamines. However, these unidentified substances were removed by selective adsorption to CG50 columns and, even if present, did not interfere with measurement of catecholamine. These results suggest that this method can be generally applied to the measurement of endogenous catecholamines. *Jpn. J. Zootech. Sci.*, 56 (8): 673-679, 1985

Dopamine (DA) has a potential inhibitory activity on prolactin (PRL) release in vitro^{1,2}) and appears to control PRL secretion through the hypothalamus-pituitary portal system^{3,4}). Since the amount of catecholamine in tissues is relatively small, development of a sensitive method is really needed to measure catecholamine especially from tissues of small experimental animals.

High pressure liquid chromatography (HPLC) has a high resolution to separate a variety of substances including catecholamines. Furthermore, an electrochemical detector (ECD) is specific and sensitive enough to detect catecholamines. For application of samples onto HPLC columns, catecholamines should be crudely extracted. Alumina (Al_2O_3) columns have been commonly used for this purpose. However, selective adsorption of catecholamine onto alumina was influenced by many factors⁵). Amberlite CG50, cation exchanger, also adsorbs monoamines, their precursor amino acids and their metabolites with a high recovery. This ion-exchange resin was mainly used for chromatographic separation of catecholamines⁶⁻⁸). Recently, Amberlite CG50 was used for selective adsorption of catecholamines in human urine and rat serum samples prior to the detection of catecholamines by HPLC with fluorometry⁹)

Present address: *Ecology Division, National Grassland Research Institute, Nishinasuno, Tochigi 329-27

or ECD¹⁰⁾. Searching for a method to extract catecholamines, we found that a single step of Amberlite CG50 chromatography can do this effectively.

In this paper, we describe a simple assay method and its application to the measurement of catecholamines in various tissue samples. Catecholamines released from hypothalami into the medium during incubation were also measured for comparison with the amounts contained in the tissues. We observed that catecholamines were extracted by Amberlite CG50 with good recovery and high specificity, and that the same species of tissue catecholamines as reported by others^{5,11,12)} were detected in comparable amounts by HPLC-ECD.

Materials and Methods

Chemicals and Reagents

Amberlite CG50 (type II) was from Tokyo Organic Chemical Co. (Tokyo), licensed maker of Rohm and Haas Co.. DA, epinephrine (E) and norepinephrine (NE) were obtained from Sigma Chemical Co. (St. Louis, MO., U.S.A.). These standards were dissolved in 0.1 N HClO₄ and stored at -20°C. Water was deionized and doubly distilled. MEM alpha medium (α -MEM) was purchased from Grand Island Biological Co. (Grand Island, NY, U.S.A.). All other chemicals were of analytical reagent grade.

Activation of Amberlite CG 50

Amberlite resin (50 g) was first suspended in water, and fine particles were removed several times by decantation. The resin was then suspended in 500 ml of 2N HCl for 30 min with stirring, washed extensively with plenty of water until pH was almost neutral, suspended in 500 ml of 2N NaOH for 30 min followed by washing with water as described above, and finally equilibrated with 0.1 M NaH₂PO₄-K₂HPO₄ (pH=6.5) containing 1 mM EDTA. The activated resin was packed to a height of 5 cm in a small glass column (0.5 i.d. $\times$ 10 cm).

Separation and quantitation of catecholamine

HPLC-ECD consisted of high-pressure pump with a pressure gauge and limiter (NSP-800-500 X, Nippon Seimitsu, Tokyo), a sample injector (NVI-3001, Nippon Seimitsu), a reversed-phase column (LiChrosorb RP-18, 5 μ m, 0.4 $\times$ 25 cm, Merck, Darmstadt, Germany) and an electrochemical detector (E-308, Irika, Kyoto). Carrier buffer was constituted of 0.1 M citric acid-0.05 M ammonium acetate (pH=3.0)-2 mM EDTA-1.25 mg/l SDS, and filtered through a nitrocellulose membrane (0.2 μ m) before use. The flow rate and applied potential were set at 1 ml/min and +0.7 V, respectively. This procedure was carried out at room temperature.

Extraction of catecholamine

Fresh tissues (approximately 3 to 10 mg) were removed and homogenized in 0.5 ml of 0.5 M HClO₄ immediately after dd mice were killed by cervical dislocation. Blood was collected by heart puncture and centrifuged at 12000 $\times$ g for 5 min to obtain blood plasma. Aqueous samples (blood plasma and α -MEM media collected after incubation of hypothalamus for 120 min) were mixed with an equal vol of 0.5 M HClO₄.

by vortexing. All samples were with dihydrobenzylamine (DHBA; 1 ng in 10 μ l of the above solution) as internal standard to monitor recovery. The sample was kept at 4°C for 30 min and centrifuged at 12000 $\times$ g for 15 min. The pellet was used for protein determination by the Lowry's method. The supernatant was centrifuged at 1000 $\times$ g for 10 min to remove insoluble KClO₄ after the pH was adjusted to 6.5–7.5 with 0.4 M K₂CO₃. One-fifth vol of 0.5 M NaH₂PO₄-K₂HPO₄ (pH=6.5) containing 1 mM EDTA was added. The sample was applied onto an Amberlite CG50 column (0.5 $\times$ 5 cm). After drained off, the column was washed with 10 ml of ice-cold water and subsequently with 1.5 ml of 0.5 N HCl-1 mM EDTA and the eluate was lyophilized thoroughly. The dried sample was dissolved in 150 μ l of 0.1 N HClO₄ and centrifuged at 12000 $\times$ g for 15 min prior to injection. The supernatant (100 μ l) was injected to HPLC-ECD.

Results

DA, DHBA, E and NE were adsorbed onto Amberlite CG50 and eluted completely with 0.5 N HCl. Recoveries of these standards in the process of CG50 chromatography were virtually 100% (102.1 $\pm$ 4.4%, 108 $\pm$ 3.8%, 102.7 $\pm$ 3.8% and 100.1 $\pm$ 1.2% for DA, DHBA, E and NE, respectively, $M\pm SD$, $n=3$). In Fig. 1 (A), these standards were clearly separated into independent sharp peaks in the chromatogram. Linear dose-response relationship was observed between the amount and the peak height for each standard ($r=0.99$), as shown in Fig. 1 (B). Intra-assay vari-

Fig. 1. Chromatogram (A) and standard curves (B) for NE, E, DHBA, and DA generated in HPLC-ECD. Catecholamines dissolved in 100 μ l 0.5 N HCl were injected to HPLC-ECD. The sensitivities were 8 nA (A) and 2 nA (B), respectively.

Fig. 2. Chromatograms of catecholamines extracted by Amberlite CG 50 from various samples. To obtain the height of peaks on chromatograms, 1 μ g adrenal gland (a), 20 μ g hypothalamus (C), 350 μ g anterior pituitary (e), 50 μ g posterior pituitary (f), 500 μ g lung (g) in terms of protein weight in tissues, 500 μ l blood plasma (b) and medium in which 3 hypothalami had been cultured for 2 hr (d) were prepared as described in the text.

ance of each standard was relatively small (2.6%, 1.4%, 2.6% and 2.2% for DA, DHBA, E and NE, respectively C.V., $n=4$). Inter-assay variance was less than 3% ($n=3$). Minimal detectable amounts of standards (signal/noise > 2) were approximately 100 fmol in DA, DHBA and E, and 75 fmol in NE, respectively. Recoveries of catecholamines added to hypothalamic homogenates were more than 90% with small SD ($90.0 \pm 5.0\%$, $92.6 \pm 4.1\%$, $90.9 \pm 3.5\%$ and $91.2 \pm 2.4\%$ for DA, DHBA, E and NE, respectively, $M \pm SD$, $n=7$). In other tissue homogenates also over 90% of recoveries was resulted (data not shown).

Catecholamines in various tissues were extracted and measured according to the procedure described above. As shown in Fig. 2, each chromatogram shows clear separation of catecholamines and minimum contamination of unidentified substances which interfere with the identification and measurement of catecholamines from tissue samples. When the samples from the medium incubated with hypothalami were applied onto alumina, the eluate contained two unidentified substances. These substances made it difficult to measure catecholamines. Such impurities were, however, eliminated by passing the medium through a column of Amberlite CG 50.

Contents of catecholamines obtained from peak heights of the chromatogram were summarized in Table 1. Adrenal glands contained the largest amounts of both E and

Determination of Catecholamines by HPLC

Table 1. Endogenous catecholamine contents in various mouse tissues

Tissue	Norepinephrine	Epinephrine	Dopamine
Adrenal gland (n mol/mg protein)	7.27 ± 1.34	16.12 ± 1.66	N. D.
Blood plasma (p mol/ml plasma)	13.2 ± 1.2	6.5 ± 1.77	N. D.
Hypothalamus (p mol/mg protein)	69.2 ± 13.4	N. D.	41.8 ± 6.5
Lung (p mol/mg protein)	N. D.	N. D.	N. D.
Pituitary anterior (p mol/mg protein)	N. D.	N. D.	4.37 ± 1.04
Pituitary posterior (p mol/mg protein)	N. D.	N. D.	38.56 ± 3.99
Hypothalamic culture (p mol/mg protein)	13.4 ± 1.77	N. D.	21.5 ± 1.36
M ± S. D. (n=3)	N. D. <75 f mol	N. D. <100 f mol	N. D. <100 f mol

NE among tissues tested. On the contrary none of these catecholamines was detected in the lung. Anterior and posterior pituitary glands had DA alone. Posterior pituitary glands (neural lobes) contained larger amounts of DA than anterior glands. The hypothalamus was rich in NE but actively secreted DA into culture medium. E was detected only in adrenal glands.

Discussion

Alumina has been widely used to extract catecholamines by many investigators^{5, 13, 14}. Our data show that Amberlite CG50 also adsorbed catecholamines selectively. The recoveries of catecholamine in the process of extraction from tissue homogenate were more than 90%, in agreement with SEKI⁷) and JACKMAN⁹). Alumina adsorbs non-catecholamine substances^{5, 14}). Such substances may interfere with measurement of catecholamine by HPLC-ECD. We also observed that unidentified substances were extracted with catecholamines from blood samples by alumina and disturbed measurement of catecholamine as described in Kato *et al.*¹⁰) in medium samples, two unidentified substances which were adsorbed onto alumina were not adsorbed onto Amberlite CG50. The HPLC-ECD has enough sensitivity to measure endogenous catecholamines, being almost comparable to radioenzymatic procedure¹⁵), gas chromatography with mass spectrometry¹⁶) or HPLC with fluorescence detection¹⁷).

Endogenous catecholamines summarized in Table 1 were compared below to the figures reported by others. Amounts of E and NE in adrenal glands were in good agreement with Anton & Sayre⁵). They found that small amounts of these catecholamines existed in the lung. We could not detect either E or NE, presumably due to the difference in the sensitivity of assay methods. The hypothalamus contained catecholamines of the same kind and of comparable amounts to those in the hypothalamus of female Snell mice and rats^{12, 18}). In the superfused rat hypothalamus, almost equal moles of NE and DA were secreted¹⁹). Our data that DA in posterior pituitary

glands is much higher than in anterior glands also agreed with those of BEN-JONATHAN & PORTER¹²⁾. The DA content in rat pituitary glands^{12,20)} are even lower than our data in mice. Some unidentified peaks were observed on the chromatogram of blood samples. These substances may be different from the ones in medium samples because they cannot be removed by the use of CG 50 resin. Even though these unidentified substances are present, NE, E and DA in blood samples are measurable by the HPLC-ECD system because the retention times of these unidentified substances are different from those of NE, E and DA. The molar ratio of NE/E in blood plasma was almost identical, while, amounts of catecholamines in plasma greatly differed depending on investigators^{5,21,22)}. Our data indicate that the extraction of tissue catecholamine by Amberlite CG50 has good recovery and specificity, and that HPLC-ECD is a suitable system for accurate determination. We conclude, therefore, that our procedures are applicable for measurement of low levels of catecholamine of small samples from mouse tissues.

References

- 1) MACLEOD, R.H. and J.E. LEHMEYER, *Endocrinology* **94**: 1077-1084. 1974.
- 2) SHAAR, C. J. and J.A. CLEMENS, *Endocrinology* **95**: 1202-1212. 1974.
- 3) TAKAHARA, J., A. ARIMURA and A.V. SCHALLY, *Endocrinology* **95**: 462-465. 1974.
- 4) GIBBS, D.M. and J.D. NEILL, *Endocrinology* **102**: 1895-1900. 1978.
- 5) ANTON, A.H. and D.F. SAYRE, *J Pharmacol Exp Ther* **138**: 360-375. 1962.
- 6) MINARD, F.N. and D.S. GRANT, *Biochem Med* **6**: 46-52. 1972.
- 7) SEKI, T., *J Chromatogr* **124**: 411-414. 1976.
- 8) SEKI, T., *J Chromatogr* **155**: 415-420. 1978.
- 9) JACKMAN, G.P., *Clin Chem* **26**: 1623-1624. 1980.
- 10) KATO, T., K. KOSHIYA and T. NAGATSU, *Protein Nucleic Acid Enzyme* **26**, 1129-1134. 1981. (in Japanese)
- 11) KISSINGER, P.T., C.J. REFSHAUGE, R. DRELLING and R.N. ADAMS, *Anal Lett* **6**: 465-477. 1973.
- 12) Ben-JONATHAN, N. and J.C. PORTER, *Endocrinology* **98**: 1497-1507. 1976.
- 13) REFSHAUGE, C.J., P.T. KISSINGER, L. DREINIGN, L. BLANK, R. FREEMAN and R.N. ADAMS, *Life Sci* **14**: 311-322. 1974.
- 14) MARUYAMA, Y. and M. KUSAKA, *Life Sci* **23**: 1603-1608. 1978.
- 15) PEULER, J.D. and G.A. JOHNSON, *Life Sci* **21**: 625-636. 1977.
- 16) JACOB, K., W. VOGT, M. KNEDEL and G. SCHWERTFEGER, *J. Chromatogr* **146**: 221-226. 1978.
- 17) YUI, Y., M. KIMURA, Y. ITOKAWA and C. KAWAI, *J Chromatogr* **177**: 376-379. 1979.
- 18) MORGAN, W.W., A. BARTKE and K. PFEIL, *Endocrinology* **109**: 2069-2075. 1981.
- 19) FOREMAN, M.M. and J.C. PORTER, *Endocrinology* **108**: 800-804. 1981.
- 20) CHIOCCIO, S.R., S. CHAFUEN and J.H. TRAMEZZANI, *Endocrinology* **106**: 1682-1685. 1980.
- 21) BUHLER, H.U., M. DA PRADA, W. HAEPFELY and G.B. PICOTTI, *J Physiol* **276**: 311-320. 1978.
- 22) BAUCE, L., J.A. THORNHILL, K.E. COOPER and W.L. VEALE, *Life Sci* **27**: 1921-1928. 1980.

イオン交換クロマトグラフィーによるカテコールアミンの抽出と 高速液体クロマトグラフィーによる測定

土肥宏志*・酒井仙吉・河本 馨・正田陽一

東京大学農学部, 東京都文京区 113

マウスの生体試料中のノルエピネフリン (NE), シビネフリン (E) およびドーパミン (DA) は, アンバーライト CG 50 により, 90% 以上の高回収率で抽出された. 抽出されたカテコールアミンは, 逆相系のカラムを用いた高速液体クロマトグラフィーと電気化学的検出器を組み合わせた方法 (HPLC-ECD 法) により分離され, クロマトグラム上のピークの高さから, カテコールアミン量を同時に測定することができた. ピークの高さと量の間には, 高い相関 ($r=0.99$) があり, 測定内および測定間変動係数は, それぞれ 2.6% と 3.0% 以下であった. また, 最低検出量は, NE が 75 fmol, E および DA が 100

fmol となり, 測定感度は鋭敏であった. さまざまなマウスの生体試料において, カテコールアミンの測定を妨害する物質は, アンバーライト CG 50 により除去されており, クロマトグラム上でカテコールアミンのピークを計測することができた.

以上の結果から, アンバーライト CG 50 と HPLC-ECD 法を組み合わせることにより, マウス生体試料中の微量なカテコールアミンを, 正確かつ同時に測定できることが明らかになった.

日畜会報, 56 (8): 673-679, 1985

*現: 草地試験場 生態部, 栃木県西那須野町 329-27