From Japan to the World

Ensuring the Reliability of Measurements of Exposure of Workers and Public to High-Energy γ -rays

—Establishment of the High Energy γ -ray Calibration Field—

Japan Atomic Energy Agency*¹

The 6–7 MeV high energy γ -ray calibration field was established at the Facility of Radiation Standards and serves to ensure the reliability of readings of dosimeters used in higher energy γ -ray fields.

The reliability dosimeter readings is ensured by so-called “calibration” which donates the periodic check of whether dosimeter readings are correct. Well-defined radiation fields called “calibration fields” are required for proper calibration. The FRS has established various kinds of radiation calibration fields including X-ray, γ -ray, β -ray and neutrons, which it offers to companies and institutes working in the field of radiation protection.

For exposure to γ -rays, fixed γ -ray calibration fields are only available up to 2 MeV, and dosimeters have been calibrated in these fields. In contrast, the emission of 6 MeV γ -rays resulting from the $^{16}\text{O}(n, p)^{16}\text{N}$ reaction in the coolant water is widely recognized as a substantial source of radiation exposure, which is particularly relevant for BWRs. Furthermore, for therapeutic and industrial purposes, γ -rays with energies greater than several tenths of MeVs are produced by electron accelerators and high energy X-ray generators. The calibration field for 6–7 MeV γ -rays should tell us how widely available dosimeters respond to γ -rays these energies or higher.

However, no high energy γ -ray calibration field is available in Japan for radiation protection. We thus have developed a 6–7 MeV γ -ray field produced by the nuclear reaction of $^{19}\text{F}(p, \alpha)^{16}\text{O}$ using the 4 MV Van-de-graaff accelerator at the FRS (**Fig. 1**) for calibration purposes.

Because of high energy of incident γ -rays, introducing an ionization chamber (IC) with a thicker build-up cap is required to precisely determine air kerma rates at a point of test (100 cm). This means that measurement conditions significantly differ from those of routine calibrations. We investigated the method of using a conventional IC coupled with the same build-up plate (BUP) introduced for routine calibrations. Throughout a series of Monte Carlo calculations, as shown in **Fig. 2**, we assessed the measurement arrangement for the conventional IC and for the BUP, and found the optimal arrangement to be a BUP made of polymethylmethacrylate with the dimension of $30 \times 30 \times 3 \text{ cm}^3$ and positioned 15 cm in front of a point of test. The method would allow us to properly calibrate the dosimeters that are most widely available.

Fig. 1 Schematic of the high energy γ -ray calibration field (at Facility of Radiation Standards).

Fig. 2 Exposures calculated at test point as a function of the thickness and shape of build-up materials.

Ideal exposures were calculated using the γ -ray fluence at 100 cm and the mass-energy absorption coefficient for dry air, which reflects the attenuation of γ -ray fluence with increasing thicknesses of build-up materials. The ideal exposure represents the exposure predicted when electrons due to γ -rays fully contribute to charges collected at a point of test as well.

In addition, introducing a cylindrical 2"φ × 2" NaI(Tl) scintillation detector as monitor allows us to quite precisely determine reference air kerma rates at a point of test when irradiating with γ -rays for regular calibrations.

REFERENCE

- 1) M. KOWATARI et al.; Measurement of air kerma rates for 6- to 7-MeV high-energy γ -ray field by ionisation chamber and build-up plate, *Radiat. Prot. Dosim.*, **162** (4), 446–458 (2014).

* This article was reprinted from “JAEA R&D Review 2014”.
<<http://jolisfukyu.tokai-sc.jaea.go.jp/fukyu/mirai-en/2014/>>