the ventricle was approximated by a cone of revolution. Twenty-three patients with normal and 23 patients with abnormal left ventricles were studied. The Dodge method provided reference values. There was no significant difference between values for EDV calculated by either method; values for ESV calculated by the cone method were consistently higher, but the error is acceptable. The cone method provides results similar to those obtained by the corrected Dodge method, eliminating the problem of choosing and measuring the long axis.

21322, 21441, 21462

Direct Serial Magnification Angiography Applied to Cerebral Aneurysms.

Zenji Shiozawa, et al., Radiology, 126(1), 263-266, (1978).

Summary

High-resolution radiographic images may be obtained by using direct serial $4\times$ angiographic magnification with a microfocus 0.05 mm x-ray tube. This technique provides more highly detailed views of cerebral aneurysms than does normal angiography, while lowering the total radiation dosage.

21322

Some Aspects of Contralateral Compression in Carotid Angiography.

Anders Möller, et al., Radiology, 125(3), 725-729, (1977).

Summary

The effect of compression of the noninjected carotid artery at carotid angiography was studied; compression above the bifurcation was compared to compression below the bifurcation. The timing of the compression in relation to the contrast injection was varied. In 12 patients an increased contrast dose was used. Retrograde contrast filling of the noninjected carotid was better when compression was applied below the bifurcation. With increased contrast dose and prolonged injection time, visualization of the noninjected carotid artery was further improved, as was contrast filling of the noninjected hemisphere.

21324, 21472

Simplified Outpatient Lower Extremity Venography.

Helene Pavlov, et al., Radiology, 126, 525-527, (1978).

23390

Studies of Pulmonary Vasculature Using Chromato-roentgenography.

Yoshiaki Tanaka and Shinji Takahashi, Radiology, 126, 527-530, (1978).

21437, 41321

CT Determination of Parameters for Inhomogeneity Corrections in Radiation Therapy of the Esophagus.

Gary D. Fullerton, et al., Radiology, 126, 167-171, (1978).

21437

Computed tomography: an overview.

Euclin Seeram, Radiologic Technology, 49(4), 491-496, (1978).

21437

Split-Detector Computed Tomography: A Preliminary Report.

Rodney A. Brooks and Giovanni Di Chiro, Radiology, 126, 255-257, (1978).

22510

Performance Characteristics of Selected Integrating Ion Chambers.

Lubenau, J. O. and Liberace R., Health Phys., 33, 199-204, (1977).

22532, 42532

Long Term Recycling Characteristics of LiF(TLD-100) Dosimeter Material.

Wald J., DeWerd, L. A. and Stoebe, T. G., Health Phys., 33, 303-310, (1977).

22532

The Energy and Directional Response of Harshaw TLD-100 Thermoluminescent Dosimeters in the Diagnostic X-ray Energy Range.

Morgan, T. J. and Bratemen, Health Phys., 33, 339-342, (1977).

224

Isodose Curves in a Phantom due to Diagnostic Quality X-Radiation.

Dale Trout E., Health Phys., 33, 359-367, (1977).

24330, 21322, 21122

A Wedge Filter for Craniocervical Angiography.

L. D. Cromwell, et al., Amer. J. Roentgenol., 129