文 献 紹 介

文献委員会

21120

Elliptical Narrow Beam Rotation Radiography.

K.A. Sämfors, et al., *Acta Radiologica Diagnosis*.
19(1), 129-138, (1978).

Summary

The basic mathematics for elliptical narrow beam rotation radiography are deduced and exemplified. An analysis of the functioning principle of elliptical systems shows that the essential factor in the image formation is the form of the evolute of the ellipse, which determines the projection of the object. The image layer does not necessarily coincide with the ellipse.

21221

Normal Movements of the Cervical Spine-Penning.

Am. J. Roentgenol. 130, 317-326, (1978)

Summary

This paper describes a technique for analyzing movement of the cervical spine. The method consists of superimposition of two films representing the cervical spine in the end positions of the movements under investigation (e.g., flexion and extension). From tracings of selected structures, movement is represented in the form of movement diagrams. Knowledge of cervical spine dynamics is helpful in understanding muscle and ligament function as well as the shape of components in various postures.

21251, 21410

350 kVp Chest Radiography: Review and Comparison with 120 kVp.

Anthony V. Proto et al., *Am. J. Roentgenol.*
130(5), 859-866, (1978).

Summary

High kilovoltage radiography had its beginning with supervoltage technique and is presently accomplished with the 350 kV chest x-ray system. The physical reasons to explain the improved visibility with this technique are: (1) absorption coefficients of bone and soft tissue, (2) more uniform bone visibility, (3) x-ray spectrum, (4) visual re-

sponse, and (5) depth resolution. In this study, 350 kV and 120 kV chest x-ray techniques were compared by evaluating 15 parameters. Based on our data and the results of previous large clinical studies, a list of disadvantages and advantages is offered. The entrance dose for an average 350 kV posteroanterior chest radiograph is 8 mR, or about one-third that with 90 kV technique.

21262, 24111

An Experimental "Trans-Molybdenum" Tube for Mammography.

G. Allan Johnson et al., *Radiology*. 127(2), 511-516, (1978).

Summary

Possible mammographic advantages of "trans-molybdenum" anodes (atomic number $>Z_{Mo}$) are decreased dose because the fluorescent radiation is more penetrating, and increased useful output; contrast degradation is known to be tolerable. The output per mAs, the HVT in Al, and the penetration in Lucite were measured spectroscopically for an experimental Rh-anode tube and also for Mo- and W-anode mammographic tubes. The trans-molybdenum tube was shown to have output and dose advantages over Mo anodes, and output and contrast advantages over normal and selectively filtered W anodes. Possible applications in areas other than mammography are briefly discussed.

21320

Fluoroscopic Guidance in Femoral Artery Puncture.

Charles T. Dotter et al., *Radiology*. 127(1), 266-267, (1978).

Summary

Analysis of 100 femoral arteriograms demonstrated a consistent relationship between the course of the femoral artery and the femoral head. Fluoroscopic guidance can be of value in femoral puncture when the arterial pulse is diminished or obscured, for training purposes, and for analysis and correction of missed punctures.

21332, 21335