41330, 411

Postoperative Radiation Therapy in Mammary Carcinoma Stage II. (Target Volume, Organs at Risk, Absorbed Dose, Time-Dose Schedule and Dose to Organs at Risk in a Prospective Investigation.)

Arwidi A. et al., Acta. Radiol (blue), 18(4), 273-281, (1980).

422, 42510, 130

A Theoretical Study of Cavity Chamber Correction Factors for Photon Beam Absorbed Dose Determination.

McEwan AC., Phys. Med. Biol., 25(1), 39-50, (1980).

summary

Theoretical expressions have been derived for the contributions to ionisation density in cavity chamber from secondary electrons arising from photon absorption in wall, build-up cap and surrounding medium. The method allows the derivation of C_{λ} and displacement factors, which are evaluated for the Farmer graphite chamber for a range of photon beams. Differences between C_{λ} values calculated for air-walled and water walled chambers are found to be consistent with those determined by Nahum and Greening. The currently recommended displacement factor for the Farmer chamber with cap, however, is found to be in error by about 0.8%. On the basis of a comparison between theoretically predicted air doses in cavities and the experimental results of Fregene it is proposed that no further correction to C_{λ} is required for the chamber effective point of measurement.

42520, 421

Ferrous Sulphate Dosimetry for Electrons. (A re-

evaluation)

Svensson H., et al., Acta. Radiol (blue), 18(4), 326-336, (1980).

summary

A theory was developed for the perturbation correction to be applied in the calorimetric determination of the ferrous sulphate dosimeter G -value. Recent determinations of the molar extinction coefficient, ϵ_m , have shown that systematic errors in ϵ_m may have influenced experimental G -values. With precision spectrophotometers this error is cancelled in the product $\epsilon_m \cdot G$. It was shown that after applying the perturbation correction a very small spread was obtained in calorimetric $\epsilon_m \cdot G$ -values, the standard error of the mean being 0.2 per cent. The mean G -value is 2.3 per cent lower than that recommended in ICRU Report No. 21.

435, 424

Accuracy of Computerized Radiation Treatment Planning.

Väyrynen T., et al., Acta. Radiol (blue), 18(4), 337-342, (1980).

summary

The accuracy of computerized dose planning depends essentially on the computer program. The quickest and simplest way to assess the accuracy of a commercial program is to compare the results of the computer calculations with measured values. The present report deals with the accuracy of the RADPLAN dose planning program in different situations easy to control. The test methods are also applicable to other corresponding systems.