scattered back into the beam through the edge of the collimator. It is demonstrated the mean energy of these electrons is only 40 per cent of the mean energy in the beam and that the smallest electron contamination is obtained if the material closest to the beam is of high density.

424, 41250

Dose Distribution Around Radium Arrays Used in the Treatment of Uterine Carcinoma.

Sharma PD, et al., *Acta Radiol.*, 19(3), 229-233, (1980).

Summary

Measurements have been made to determine the dose distribution around typical arrays of commercially available Campbell type Heyman applicators using thermoluminescent dosimeters. These measurements verify the tables produced by Benner, linking the number of sources used in the so called packing technique for the treatment of uterine carcinoma to an optimum treatment time.

424

A Simple Function Describing the Absorption in Platinum for Dose-Rate Calculations Around Radium Applicators.

Toepfer KD, et al., *Brit. J. Radiol.*, 53, 1078-1082, (1980).

Summary

The diminution of exposure rate from radium γ radiation by filtration through platinum can be described very accurately by $T_0/(1+\mu_0 d)$ for wall thicknesses d from 0.2 mm to 4.0 mm ($T_0=6.356 \times 10^{-5}$ Am²/kg² or 8.869 R cm²h⁻¹mg⁻¹; $\mu_0=1.501$ cm⁻¹). This function fits generally accepted data to within experimental error, to an accuracy not

achieved by any single exponential factor. The expression is derived theoretically from a simple approximation to the spectral distribution of radium γ radiation.

With this type of transmission function, the expression for the exposure rate around linear radium sources in platinum tube (generalized Sievert integral) is readily integrated in terms of elementary functions. Used in place of the Sievert integral in calculations of dose distribution, this new formulation makes possible straight forward numeral and avoids the storage of tabulated values in computer programmes.

42532

Pion Dosimetry with Thermoluminescent Materials.

Hoffmann W, et al., *Phys. Med. Biol.*, 25(5), 913-921, (1980).

51310

医療被曝のリスク推定方法について

橋詰 雅他, *日本医放会誌*, 40(12), 1175-1182, (1980).

610, 21437

Accuracy of Echorenticulography as Compared with Computer Tomography in Children.

Evasmie U, et al., *Acta Radiol. (red)*, 21, 315-320, (1980).

610

Ultrasonic Propagation Properties of Collagen.

Goss SA, et al., *Phys. Med. Biol.*, 25(5), 827-837, (1980).

620

An Improved Computer/Thermograph Interface.

Newman P, *Phys. Med. Biol.*, 25(5), 933-934, (1980).