文 献 紹 介

24520, 24263

A Physical Assessment of the Imaging Performance of a Panel-Type X-ray Image Intensifier.

Hay GA, et al., British J. Radiol., 54, 24~28, (1981).

Summary

The performance of a commercial panel-type X-ray image intensifier has been investigated by physical methods, both objective and subjective. The objective measurements show that the manufacturer's claims are justified, except perhaps in regard to contrast loss. The subjective measurements reveal that the conversion factor ("gain") of the device is such that dark-adaptation is not necessary, through best results are obtained in a nearly dark room. Hence the device confers a significant advantage, both in information and in dose to the patient, over traditional fluoroscopy. However, the conversion factor is nearly three orders of magnitude smaller than that of a conventional X-ray image-intensifier television system. Therefore, at modern fluoroscopic exposure rates, the eye is operating at light levels at which its own deficiencies set a limit of the perceptible information. To equal the performance of modern fluoroscopy the exposure rate to the intensifier must be increased to about 20 times the value customary in modern fluoroscopy. This can be done only by increasing the X-ray factors (kV and mA), resulting in a corresponding increase in dose to the patient of about one order of magnitude. Such an increase is in many circumstances unacceptable.

24540, 24550, 21320

Angiography Assisted by a New Videostore system.

Dossetor RS, British J. Radiol., 54, 48-50, (1981).

350, 32234

ROC Analysis of Diagnostic Performance in Liver Scintigraphy.

Fritz SL, et al., J. Nucl. Med., 22(2), 121-128, (1981).

Summary

Studies on the accuracy of liver scintigraphy for the detection of metastases were assembled from 38 sources in the medical literature. An ROC curve was

fitted to the observed values of sensitivity and specificity using an algorithm developed by Ogilvie and Creelman. This ROC curve fitted the data better than average sensitivity and specificity values in each of four subsets of the data. For the subset dealing with Tc-99m sulfur colloid scintigraphy, performed for detection of suspected metastases and containing data on 2,800 scans from 17 independent series, it was not possible to reject the hypothesis that interobserver variation was entirely due to the use of different decision thresholds by the reporting clinicians. Thus the ROC curve obtained is a reasonable baseline estimate of the performance potentially achievable in today's clinical setting. Comparison of new reports with these data is possible, but is limited by the small sample sizes in most reported series.

34510

Effects of Dissolved Oxygen of Tc-99m Methylene Diphosphonate: Concise Communication.

Coupal JJ, et al., J. Nucl. Med., 22(2), 153-156, (1981).

34130

A Comparison of Count Rate Parameters in Gamma Cameras.

Lewellen TK, et al., J. Nucl. Med., 22(2), 161-168, (1981).

Summary

Pulse pile-up is a fundamental problem that limits the ability of a gamma camera to produce high quality images at high count rates. Pulse pile-up results in loss of events and spatial distortions in the image. The question asked in this study is how well do some of the test procedures for measuring count rate performance in gamma cameras compare. In four gamma cameras we compared measurements of spectral fraction, count-rate curve, pulse-pair resolution, deadtime, maximum count rate, full width at half maximum of the line spread function, and misplaced event count rate. The results indicated that no one technique provides a complete description of the count-rate effects in gamma cameras. The misplaced-event measurements provided the most information.