In common spondylolysis occur frequently in the L4-5, L5-S area, I performed attesting in order to improve imaging radiograph in making the direction centerray angle cranial 30° after making the patient in the supine position. As the result of it, both pars interarticularis imaging appeared in the more superior reappearance than it appeared in other radiographs.

So I am reporting.

(6) Measurement of Leg Length by CT Scanogram and Orthoroentgenogram

Ladda Chaloeykitti B. Sc., Laddawan Treeyapattanporn B. Sc., Piya Netrvichian, M.D.

Department of Radiology, Faculty of Medicine, Chiang Mai University, Kingdum of Thailand.

Abstract

The technique has been used to evaluated the leg length discrepancy is called orthoroentgenogram. This method, the three exposures centering over hips, knees and ankles are made on a single film with the radiopaque ruler beneath the patient.

Recently the use of computed tomographic scanogram was introduced with more accuracy, more simple and reduction of radiation dosage. In our study, compared the two methods, in the accuracy, the radiation dose, the time and the cost.

3. QC & PACS

Chairman Kazuya Yamaguchi

(7) Computerized Imaging Management Using LAN in Department of Diagnostic Radiology

Hwan Cheung, Sung-Ryong Kim

Department of Diagnostic Radiology, Seoul National University Hospital, Republic of Korea.

Abstract

Using Local Area Network, archiving and communication of radiological image are managed systematically. Conventionally, after radiographic study, films and reports were transmitted to ward or diagnostic rooms manually but this system make it unnecessary and reduce the requirement of manpower and delay time significantly. Therefore, in ward or diagnostic rooms, whenever radiological informations, such as image and report, are necessary, they can be accessed directly using this system. Besides, this system will be used as the basic medical information system through which, hospital with poor radiographic unit can share the capacity of large hospital and general level of health care be improved remarkably.

(8) Quality Control in a Mammography Environment

Mary Lou Dysart

Adoministrator of Radiology, Beth Israel Medical Center, U.S.A.