学術交流委員会だより

平成12年度後期国際研究集会派遣会員報告書

派遣員 吉田宣博 愛知県厚生連海南病院

派遣国際研究集会：86th Scientific Assembly and Annual Meeting of the Radiological Society of North America (RSNA)

開催場所：McCormick Place, Chicago, Illinois, U.S.A

開催期間：Nov. 26-Dec. 1, 2000

Evaluation of Cardiac Function by ECG-Gated SPECT Using the QGS program: Evaluation by the Dual Isotope Method ($^{201}\text{Tl} + ^{123}\text{I}$)

Yoshida N

Purpose:

It is important to obtain accurate information about myocardial blood flow, metabolism, and function for evaluating the pathologic conditions and therapeutic values in various cardiac diseases. However, these 3 parameters can be simultaneously obtained at the same site when dual SPECT ($^{201}\text{Tl} + ^{123}\text{I}$) is combined with the QGS program. However, it was reported that dual SPECT is not appropriate for quantitative evaluation because of the influence of cross-talk. Therefore, we evaluated the influence of cross-talk on QGS.

Methods:

A cardiac phantom (RH-2 type) was separately labeled with ^{201}Tl and ^{123}I , and SPECT images were acquired three times. In addition, the same cardiac phantom was simultaneously labeled with both ^{201}Tl and ^{123}I , and SPECT images were acquired three times as well. The conditions of image acquisition were as follows: the energy window for ^{201}Tl was established at 68 keV (20%), whereas the energy window for ^{123}I was established at 159 keV (20%). SPECT images were acquired at photopeaks of the respective isotopes.

Results:

1. Phantom: When the volumes obtained by the single isotope method were compared to those obtained by dual SPECT, the volumes obtained by dual SPECT were influenced by cross-talk. That is, the volume obtained by the count levels of ^{123}I was underestimated by 3.4%, while the volume obtained by the count levels of ^{201}Tl was underestimated by 5.5%.

2. Clinical cases: (1) When the volumes obtained by the single isotope method were compared to those obtained by the dual SPECT, the volumes obtained by the dual isotope method were influenced by cross-talk. That is, the volumes obtained by the count levels of ^{123}I were underestimated by 5.0% during the diastole and by 24.1% during the systole. Moreover, the volumes obtained by the count levels of ^{201}Tl were underestimated by 8.6% during the diastole and by 20.9% during the systole. (2) When the ejection fraction (EF) obtained by the single isotope method was compared to that obtained by the dual isotope method, EF obtained by dual SPECT was overestimated by 9.8% on the basis of count levels of ^{123}I and by 5.3% on the basis of count levels of ^{201}Tl .

Conclusion:

Cross-talk greatly influenced not only the volumes calculated by QGS, but also EF. Therefore, QGS analysis for dual SPECT is considered problematic.

McCormick Place, Chicago, Illinois, U.S.Aにて86th Scientific Assembly and Annual Meeting of the Radiological Society of North America (RSNA)がNovember 26~December 1, 2000の期間にて開催されました。私は、以前よりRSNAに参加したいと考えていましたが、今回初めて申し込んだ演題が採用され、この世界最大の放射線学会に参加できたことは幸運だと思っています。会場になったMcCormick Placeは、ミシガン湖のほとりにありシカゴ市内から5kmほど南に位置しLake Side Building, South Building, North BuildingからなりSouth BuildingとNorth BuildingのLEVEL 3では、Technical Exhibits(機器展示)が開催され、広大な会場に600社以上もの企業の最先端の医療機器が展示してありました(Photo. 1)。なかでも興味深く見学したのは、日本放射線技術学会では展示のないシングルフォトンECT(SPECT)の検出器やポジトロン・エミッション・トモグラフィ(PET)の検出器であり、可変型の検出器や外部線源を用いて吸収補正するもの、またはPETやSPECTの検出器とCTが一緒になっているものなど初めて見ることができ、非常に勉強になりました。

学会期間中の講演や演題数は、Refresher Courses：275題、Scientific Paper Sessions(口述発表)：1,672題、Scientific Poster Sessions(ポスター展示：技術的)：406題、Education Exhibits(ポスター展示：学識的)：990題と多数あり、私の発表区分は、Scientific Poster SessionsのNuclear Medicineでした。演題は、「Evaluation of Cardiac Function by ECG-Gated SPECT using the QGS program -Evaluation by the Dual Isotope Method ($^{201}\text{Tl} + ^{123}\text{I}$)-」で、日本では、最近注目を浴びている心筋SPECTの解析ソフトであるquantitative gated SPECT(QGS)について検討したものです(Photo.

Photo. 1 North Buildingの機器展示場の前で.

2). 不思議に感じたのは日本の放射線学会では、このQGSの検討は数多く発表されていますが今回のRSNAでは、私の演題一題であったことです。アメリカでは、日本より早くこのソフトが導入されており、すでに十分な検討がなされているせいかもしれませんが、日本とアメリカの違いを感じました。

私の発表内容は、2核種同時収集($^{201}\text{Tl} + ^{123}\text{I}$)で行った場合、cross talkがQGSの解析値にどの程度影響を与えるかを定量的に検討したものです。要旨を以下に示します。

「心疾患の病態把握や治療効果の評価には、心筋の血流、代謝、機能情報を正確に把握することが重要になるが、dual SPECT($^{201}\text{Tl} + ^{123}\text{I}$ -BMIPP)とQGSプログラムを組み合わせることによりこの三つのパラメータを同時期、同位置で求めることができ臨床上有用である。しかし、dual SPECTは、cross talkにより定量的評価法には適切でないとの報告もある。そこで今回は、cross talkがQGSに与える影響について検討した結果、単独核種よりdual SPECTの容積の方が低く算出され ^{123}I 、 ^{201}Tl それぞれ3.4%、5.5%の過小評価となった。また臨床例において駆出率を求めた結果、dualで撮像した方が ^{123}I では7.2%、 ^{201}Tl では5.2%の過大評価

Photo. 2 自分の発表のポスター.

となった。cross talkの影響は、容積のみならず駆出率にも大きく、dual SPECTにおけるQGSの解析には問題があると考えられる」詳細は、英文abstractを参照してください。他の発表を見るとほとんどが医師であり放射線技師の発表は少ないように感じました。みなさんどんどん応募してみたいかがでしょうか。

RSNAに参加し全体を通して好感をもてた点は、シャトルバスが常時運行しており学会に行きやすい点、会場内には、いくつかのHelp Cornerが設置してあり対応も親切であり参加者に対する配慮が十分感じられた点でした。また会員、非会員にかかわらず演題応募可能なことや発表者の参加登録費が無料である点など誰でも発表しやすいシステムになっており、すばらしいと感じました。

最後に、今回のRSNA2000の参加に対し日本放射線技術学会より国際研究集会派遣会員としてとして貴重な経験をさせていただきました。川村学会長をはじめ関係者の方々には、深く感謝いたします。