CP7：総合企画 7 「New technology が創造する泌尿器科の近未来」

司会のことば

関西医科大学¹⁾，北里大学²⁾

松田公志¹⁾，馬場志郎²⁾

医療の発展は、医学による真実の解明とともに Technology の進歩によって支えられてきた。New Technology が疾患の治療法を根本的に変革する最も顕著な例を、われわれ泌尿器科医は尿路結石症に対する ESWL の導入で目の当たりにした。今、さらに加速度的な Technology の進歩が医療を変えつつある。

わが国においても手術用ロボットが導入され、泌尿器科でもすでに実際に使用している施設もある。遠隔手術も行われた。こうした技術は、従来の手術に比べてどのような長所を持っているのか、泌尿器科の未来を大きく変えるパワーがあるのか。New Technology と泌尿器科を考えると、代表的なテーマといえよう。超音波駆動メスやシーリングシステムの導入は、日常の手術を大きく変革したが、今後どのようなエネルギーソースが開発されるのか、新しい画像診断の開発状況とともに、泌尿器科医にとって大きな関心事である。さらに、カプセル型胃カメラが実用化されつつあるように、ナノテクノロジーが医療にもたらすインパクトは計り知れないものがある。

Technology の最先端はどこまで来ているのか、そしてわれわれ泌尿器科医は、それをどのように利用できるのか。与えられた完成品を使うだけでなく、臨床への有効利用法を提案するのもわれわれ臨床医の大きな役割といえよう。

CP7-1

Robotic surgery for urology in Japan

Keio University

Ken Nakagawa, Masaru Murai

Laparoscopic surgery in urology has developed for the last decade as well as other surgical fields. But laparoscopic surgery has not become a common technique, because the inherent limitations of traditional laparoscopic surgery, the two-dimensional laparoscopic view and the limited motion of straight laparoscopic instruments through a trocar, make it difficult to change from open surgery. The robotics may be a solution for the difficulty. Two telerobotic systems are currently commercially available, Zeus (Computer Motion) and da Vinci (Intuitive Surgical). These systems are in the beginning stage in Japan. In the urologic field, preliminary clinical trials have been started. We tried da Vinci for seven cases of retroperitoneoscopic radical prostatectomies. Da Vinci consists of three parts, a working console, a video equipment tower and a surgical arm cart. Da Vinci is characterized by the three dimensional imaging system and the manipulator arm with six degrees of freedom. Our initial experience of retroperitoneoscopic radical prostatectomies with da Vinci was successfully accomplished, excepting first case that took long operating time. The average operating time was 237 min (range, 203—300 min) and blood loss ranged from 50 ml to 450 ml. The rapid learning curve was recognized in the series. The urinary catheter was removed within 5 days after the operation. The articulated instruments of da Vinci seem particularly suited for the difficult anastomosis between the urethra and bladder and they supplied shortened suturing time and superior anastomosis. The feeling of the handling was completely new and excellent, comparing with traditional laparoscopic surgery and open surgery. The robotic camera control provided rapidity and steadiness and the 3-D camera made the operating field wider and deeper. The radical prostatectomy using da Vinci system is feasible. But it may owe to experienced laparoscopic surgeons because the strong robotic arms lack tensile feedback.