A Simplified Radioimmunoassay of Adenosine-3':5'-Monophosphate (Using Dextran-Coated Charcoal)

Yoshiki KATO, Junichi TAKEZAWA, Morio SUZUKI,
Akira KUNINAKA and Hiroshi YOSHINO

Research Laboratory, Yamasa Shoyu Co., Ltd.,
550, 2-chome, Araoi, Choshi-shi, Chiba-ken

Received September 28, 1974

Dextran-coated charcoal was proved to be able to separate free cAMP from antibody-bound cAMP. Only free cAMP was adsorbed on dextran-coated charcoal within 1 min after contacting the charcoal and reaction mixture of cAMP and anti-cAMP-plasma, and most of antibody-bound cAMP was not adsorbed within 4 min after contacting. The data obtained were confirmed to be almost equal to the data of another experiment using cellulose ester filter separation technique. Thus, dextran-coated charcoal could be employed to simplify the radioimmunoassay of cAMP.

1. Introduction

The role of adenosine-3':5'-monophosphate (cAMP) in cellular metabolism has been a subject of intense interest in life sciences, and much effort has been devoted to the development of simple and reliable assay procedures^{1)~7)}. Of the numerous methods available for the assay of cAMP in biological samples, radioimmunoassay and competitive protein-binding method of estimation have increased in popularity because of their high sensitivity and specificity. The binding protein is relatively easy to obtain, but it is known to be more unstable than anti-serum^{4), 8), 9)}. On the other hand, the anti-cAMP serum, which seems more difficult to prepare than the binding protein, has been thought to possess greater specificity for cAMP as well as higher stability than the binding protein^{3), 4), 8), 9)}. Therefore, the radioimmunoassay is believed to be a very useful method especially for cAMP in unpurified tissue.

In the radioimmunoassay of cAMP, separation of free cAMP from antibody-bound cAMP is one of the most important processes. So far the following 3 methods have been employed for the separation: Precipitation

of antibody-bound fraction with a second antibody¹⁾, precipitation of antibody-bound fraction with ammonium sulfate²⁾, and adsorption of antibody-bound fraction on cellulose ester filters such as Millipore filter^{3), 4)}. However, these methods are not necessarily satisfactory. The double antibody radioimmunoassay procedure was reported to be time-consuming and troublesome⁴⁾. The data obtained from ammonium sulfate precipitation technique was described to vary widely from tube to tube⁴⁾. Although cellulose ester filter adsorption technique has been regarded as a simple and excellent method⁴⁾, the expensive technique was not completely reliable in our experiments: In some cases antibody-bound cAMP was not quantitatively adsorbed on the cellulose ester filter.

During investigation of the radioimmunoassay technique, we found it possible to separate free cAMP from antibody-bound cAMP by using dextran-coated charcoal. Although separation with dextran-coated charcoal has been employed for the binding protein method⁷⁾, it has not been used for radioimmunoassay of cAMP, because dextran-coated charcoal was reported to adsorb not only free cAMP but also antibody-bound

cAMP²³. We carefully studied conditions of the adsorption. As a result, it was confirmed that only free nucleotides were adsorbed on dextran-coated charcoal within 1 min after contacting but most of antibody-bound nucleotides were not adsorbed within 4 min after contacting.

Basing upon the above confirmation, we succeeded in a modification of the radioimmunoassay of cAMP. The modification would render the radioimmunoassay essentially as rapid and inexpensive as the protein-binding assay.

2. Materials and Methods

1) Synthesis of 2'-O-succinyl cAMP

2'-O-succinyl cAMP was synthesized by S. Shibuya in our laboratory according to the procedures of Steiner, et al.¹³

2) Preparation of antigen

2'-O-succinyl cAMP was coupled to human serum albumin by the method of Steiner, et al.¹³

3) Preparation of cAMP-binding antibody

Fifteen rabbits were immunized with above antigen according to the procedures of Steiner, et al.¹³ Twelve "anti-cAMP-plasma" which contained significant antibody titers (>40% binding of ³H-cAMP at 1:50 dilution) were obtained, and was used at 1:50-200 dilution with 0.8% NaCl solution or 0.05 M sodium acetate buffer, pH 6.2, containing 0.2% bovine serum albumin. In this paper, the "anti-cAMP-plasma" was employed as cAMP-antibody, because the binding activity of the plasma was confirmed to be essentially equal to that of the sera obtained from the same blood and plasma were more easily prepared in higher yields than sera.

4) ³H-cAMP

Commercial high specific activity of ³H-cAMP (30 Ci/mmole, RCC) was diluted 1:2500 with 0.8% NaCl solution or 0.05 M sodium acetate buffer, pH 6.2, and stored at -20°C in small aliquots.

5) Cyclic nucleotides and nucleotides

Cyclic AMP, guanosine-3':5'-monophos-

phate (cGMP), uridine-3':5'-monophosphate (cUMP), ATP, ADP, GTP and 5'-AMP were commercial products of our laboratory, Adenosine-2':3'-cyclic monophosphate (2':3'-cyclic AMP) was obtained from Koso Kagaku Co.

6) Other chemicals

Dextran preparations, T-40~T-250, were obtained from Pharmacia Fine Chemicals. T-40 was the most suitable for our purpose. Charcoal preparations were obtained from Hitachi Tanso Co., Ltd. Nittan SM-17 was the most available. Bovine serum albumin, fraction V, was obtained from Armour Pharmaceutical Co.

7) Preparation of scintillator solution

A toluene-based scintillator containing Triton X-100 (2:1)²⁴ was used.

3. Results

1) Kinetics of adsorption of cAMP on dextran-coated charcoal

Reaction mixtures of ³H-cAMP and anti-cAMP-plasma were mixed with dextran-coated charcoal suspension, agitated immediately, and centrifuged at regular time intervals indicated in Fig. 1. Counts of the supernatants were plotted against time. Three phases were observed: Phase I (0 to 1 min after mixing), phase II (1 to 4 min after mixing) and phase III (5 to 20 or more min after mixing). In phase I, most of free cAMP in the reaction mixtures was rapidly adsorbed on the charcoal. In phase II, most of antibody-bound cAMP seemed to remain in the supernatant. In phase III, antibody-bound cAMP was supposed to be dissociated gradually into free cAMP and antibody, and the liberated free cAMP would be adsorbed on the charcoal.* Thus, when centrifugation

* The supposition is based on the 2 facts: 1) Activity of anti-cAMP-plasma did not significantly decrease after contacting the anti-cAMP-plasma with charcoal for 90 min. 2) When inactive cAMP was incubated with anti-cAMP-plasma for 60 min at 0°C and the reaction mixture was contacted with charcoal for 0 to 90 min at 0°C, the activity of the supernatant to bind to ³H-cAMP increased as time went.

Fig. 1 Kinetics of adsorption of cAMP on dextran-coated charcoal.

Each tube contained 500 μ l of water, 200 μ l of 0.5% bovine serum albumin-contained 0.1M acetate buffer, pH 6.2, (0.5% BSA buffer), 100 μ l of 3 H-cAMP (ca. 33,500 dpm), and 100 μ l of anti-cAMP-plasma No. 4-9 (diluted 1:100 with 0.8% NaCl solution (saline)) or saline in a final volume of 900 μ l. Incubations were carried out for 60 min at 0°C. Each of reaction mixtures was mixed with 100 μ l of dextran-coated charcoal suspension (DCC), agitated immediately, and centrifuged at regular time intervals indicated in Fig. 1. Counts of supernatants were plotted against time between mixing and centrifuging.

●: anti-cAMP-plasma +; ○: anti-cAMP-plasma - (blank)

was carried out in phase II, especially between 1.2 and 1.8 min or between 2.5 and 4.0 min after mixing, counts in the supernatants were almost constant. The counts were confirmed to be almost equal to the counts of antibody-bound cAMP obtained in another experiment using membrane filter separation technique. [The counts of the supernatants between 1.2 and 1.8 min were a little higher than those between 2.5 to 4.0 min, and were closer to the values in membrane filter separation technique.]

2) Effect of pH on cAMP-antibody reaction

As shown in Fig. 2, high affinity between cAMP and antibody was observed at pH 5.5 to 6.2.

3) Effect of incubation time on cAMP-antibody reaction

A 60 min incubation was observed to be sufficient for cAMP-antibody reaction under

Fig. 2 Effect of pH on cAMP-antibody reaction.

Each tube contained 200 μ l of water, 500 μ l of 0.2% bovine serum albumin-contained 0.05 M acetate buffer (pH 4.4, 5.0, 5.6 or 6.2) or 0.05 M Tris-HCl buffer (pH 7.5), 100 μ l of 3 H-cAMP (ca. 32,000 dpm), and 100 μ l of anti-cAMP-plasma No. 4-2 (diluted 1:50 with saline) or saline in a final volume of 900 μ l. Incubations were carried out for 90 min at 0°C. Each of reaction mixtures was mixed with 100 μ l of DCC, agitated immediately, and centrifuged. Counts of supernatants were plotted against pH. All determinations were done in duplicate.

●: anti-cAMP-plasma +; ○: anti-cAMP-plasma - (blank)

Fig. 3 Effect of incubation time on cAMP-antibody reaction.

Each tube contained 250 μ l of water, 500 μ l of 0.2% bovine serum albumin-contained 0.05 M acetate buffer, pH 6.2 (0.2% BSA buffer), 100 μ l of 3 H-cAMP (ca. 35,000 dpm), and 50 μ l of anti-cAMP-plasma No. 4-5 (1:25 with saline) in a final volume of 900 μ l. Incubations were carried out for 15 to 150 min indicated in Fig. 3 at 0°C. Each of reaction mixtures was mixed with 100 μ l of DCC (which contained Nittan WSBB in place of SM17, and Nakarai Chemicals' dextran (M. W. 60,000) in place of dextran T 40), agitated immediately, and centrifuged. Counts of supernatants were plotted against incubation time. All determinations were done in duplicate.

Fig. 4 Effect of amount of anti-cAMP-plasma on cAMP-antibody reaction.

(A) Each tube contained 0 to 750 μ l of 0.5 % BSA buffer, 100 μ l of ^3H -cAMP (ca. 36,500 dpm), and 50 μ l of 0.5 % BSA buffer or 50 to 800 μ l of diluted (1:50 with 0.5 % BSA buffer) or nondiluted anti-cAMP-plasma of No. 2-9 (■), No. 9-3 (×) or No. 15-8 (○) in a final volume of 900 μ l. (B) Each tube contained 0 to 250 μ l of saline, 500 μ l of 0.2 % BSA buffer, 100 μ l of ^3H -cAMP, and 50 μ l of saline or 50 to 300 μ l of antiserum No. 4-5 (diluted 1:100 with saline) (□) in a final volume of 900 μ l. (C) Each tube contained 100 to 250 μ l of 0.5 % BSA buffer, 100 μ l of ^3H -cAMP, and 50 μ l of 0.5 % BSA buffer or 50 to 200 μ l of anti-cAMP-plasma No. 4-7 (diluted 1:20, 1:160 or 1:1280 with 0.5 % BSA buffer) (●) in a final volume of 400 μ l. Incubations were carried out for 60 min at 0°C. Each of reaction mixtures was mixed with 100 μ l of DCC, agitated immediately, and centrifuged. The blank dpm was subtracted from each result to give the dpm bound. All determinations were done in duplicate.

the examined condition (Fig. 3).

4) Effect of amount of anti-cAMP-plasma on cAMP-antibody reaction

The extent of binding of cAMP with anti-cAMP-plasma was plotted against amount of the anti-cAMP-plasma. Usually the dilution chosen is that required to bind 35 to 65 % of added ^3H -cAMP. In our preparations, the optimum amount of anti-cAMP-plasma was 20 to 0.5 μ l/tube containing ca. 0.5 pmoles of ^3H -cAMP, and the maximum binding obtained was ca. 80 % (Fig. 4).

5) Typical standard immunoassay curves for cAMP

Fig. 5 Typical standard immunoassay curves for cAMP.

Each tube contained 0 to 500 μ l of water, 200 μ l of 0.5 % BSA buffer, 0 to 500 μ l of standard cAMP solution (200, 80, 40 or 20 pmoles/ml water), 100 μ l of ^3H -cAMP (34,000 dpm), and 100 μ l of anti-cAMP-plasma No. 4-5 (diluted 1:10 (○), 1:25 (●), 1:50 (△) or 1:100 (▲) with saline) or saline (blank) in a final volume of 900 μ l. Incubations were carried out for 60 min at 0°C. Each of reaction mixtures was mixed with 100 μ l of DCC, agitated immediately, and centrifuged. The blank dpm was subtracted from each result to give the dpm bound. Counts of ^3H -cAMP bound (dpm/tube) were plotted against total cAMP; added cAMP and ^3H -cAMP (pmoles/tube). All determinations were done in duplicate.

A typical standard immunoassay curves are shown in Fig. 5. Displacement of ^3H -cAMP by unlabeled cAMP, when plotted as a semilogarithmic function, has linearity in each assay. Standard curves were converted to straight line (Fig. 6) using a simple equation according to Okabayashi, et al.⁴⁾ The lowest amount of cAMP measurable by this method was 0.5 pmole.

6) The inhibition of ^3H -cAMP binding to cAMP antibody by various nucleotides

The specificity of the antibody was tested against several nucleotides related to cAMP structurally. As shown in Fig. 7, 5'-AMP, 2':3'-cyclic AMP, ATP, ADP, cGMP, cUMP and GTP exhibited negligible cross-reactivity.

7) Standard immunoassay procedure

Basing on the above results, the following procedure was established:

Each tube contained (in order of addition) 0 to 500 μ l of water, 0 to 500 μ l of sample or standard cAMP solution, 200 μ l of

Fig. 6 A typical standard curve for cAMP (alternative plot)

Each tube contained 0 to 200 μ l of water, 500 μ l of 0.2% BSA buffer, 0 to 200 μ l of standard cAMP solution (4 to 200 pmoles/ml of water), 100 μ l of 3 H-cAMP (ca. 34,000 dpm), and 100 μ l of anti-cAMP-plasma No. 4-5 (diluted 1:100 with saline) or saline (blank) in a final volume of 900 μ l. Reaction mixtures were incubated for 60 min at 0°C, mixed with 100 μ l of DCC, agitated immediately, and centrifuged. The blank dpm was subtracted from each result to give the dpm bound in the presence of standard cAMP. C_0/C_x for each level of cAMP was calculated and plotted against amount of inactive cAMP per tube on linear graph paper. All determinations were done in duplicate.

0.5% bovine serum albumin-contained 0.1 M acetate buffer (pH 6.2), 100 μ l of 3 H-cAMP (ca. 20 nCi, ca. 0.5 pmoles) and 100 μ l of anti-cAMP-plasma in 0.8% NaCl solution or in above buffer at a dilution 1:50 to 200 sufficient to bind 35 to 65% of the labeled ligand in a final volume of 900 μ l. The reaction mixture was incubated 60 min at 0°C. A hundred μ l of dextran-coated charcoal suspension, which had been prepared by mixing 150 mg of charcoal (Nittan SM 17) and 135 mg of dextran (T-40) in 3 ml of above buffer, was added to the reaction mixture within 2 min. After brief agitation charcoal-bound cyclic nucleotide was spun down at 2°C (at 1000 g for 10 min). Radioactivity of an aliquot (500 μ l) of the supernatant containing antibody-bound cAMP was counted. It is essential to start the centrifugation between 1.2 and 1.8 min or between 2.5 to 4.0 min after mixing dextran-coated

Fig. 7 The inhibition of 3 H-cAMP binding to cAMP-antibody by various nucleotides.

Each tube contained 200 μ l of 0.5% BSA-0.1 M acetate buffer (pH 6.2), 100 μ l of 3 H-cAMP (ca. 33,000 dpm), 500 μ l of nucleotide solution (2×10^{-4} to 2×10^{-8} M in water), and 100 μ l of anti-cAMP-plasma No. 4-5 (diluted 1:100 with 0.8% NaCl solution) in a final volume of 900 μ l. Incubations were carried out for 60 min at 0°C. Each of reaction mixtures was mixed with 100 μ l of DCC, agitated immediately, and centrifuged. All determinations were done in duplicate. The blank dpm was subtracted from each result to give the dpm bound. Relative % of bound 3 H-cAMP were plotted against nucleotide concentration.

○: cAMP; ●: 2':3'-cAMP; △: 5'-AMP; ▲: ATP; ▼: ADP; □: cGMP; ■: GTP; ×: cUMP

charcoal and the reaction mixture. Under such time limit and the limit dictated by centrifuge size, as many as 24 tubes (usually, 4 levels of standard cAMP and 8 samples in duplicate) can be assayed at once, and one person can measure over 40 samples in duplicate per day.

8) Estimation of cAMP level in human plasma

Various concentrations of standard cAMP solution (0.4 ml each) were added to two kinds of human plasma (0.8 ml), whose endogenous cAMP had previously been diminished by the action of endogenous phosphodiesterase during preincubation of plasma at room temperature for a few hours, or 2% bovine serum albumin in 0.8% NaCl solution at 0°C. Cyclic AMP was recovered from plasma and BSA solution by the Somogi method at 0°C. Each sample was mixed with

Fig. 8 Calibration curves of cAMP level in human plasma.

Each tube contained 0 to 200 μ l of 0.8% NaCl, 0.2% BSA-0.05 M acetate buffer (pH 6.2), 100 μ l of 3 H-cAMP (ca. 33,000 dpm), 100 to 200 μ l of standard cAMP solution or sample treated as described in text, and 100 μ l of anti-cAMP-plasma (No. 4-5, diluted 1:100 with 0.8% NaCl solution) in a final volume of 900 μ l. Incubations were carried out for 60 min at 0°C. Reaction mixtures were mixed with 100 μ l of DCC, agitated immediately, and centrifuged. All determinations were done in duplicate.

○, ●: human plasma; △: 2% BSA;
▲: standard cAMP solution

0.4 ml of 0.3 N Ba(OH)₂. The mixture was agitated and allowed to stand at 0°C for 15 min. Then 0.4 ml of 5% ZnSO₄ was added. The mixture was agitated and centrifuged at 1000 g for 15 min at 2°C. The calibration curves were prepared by radioimmunoassay of these supernatants, and the standard curve was prepared by radioimmunoassay of standard cAMP solutions. These calibration curves were parallel with the standard curve (Fig. 8). This fact indicated that cAMP in human plasma can be determined by using the normal standard curve. The method was used to measure plasma cAMP in six normal, young adult males. The levels obtained were 16 to 26 pmoles/ml.* These values are in reasonable agreement with those in the literature.⁷⁾

* This levels were estimated twice in duplicate. Intraassay difference of cpm between two tubes was less than 5% of total count, and interassay difference was less than 10% of estimated mean value.

4. Discussion

Dextran-coated charcoal was successfully employed in radioimmunoassay of cAMP. Dextran-coated charcoal was proved to be effective to separate antibody-bound cAMP from free cAMP. Various molecular sizes of dextrans (T20~T250) tested could be used except water-insoluble T2000. Among them dextran T40 was especially suitable for this purpose, because the counts of blank were lower than others. Various kinds of charcoal tested were well suited for the purpose except Nittan H that did not well adsorb cAMP.

In our experience, several kinds of cellulose ester filters could not always adsorb antibody-bound cAMP quantitatively. On the other hand, dextran-coated charcoal could always select free cAMP and antibody-bound cAMP under the conditions we established.

Prior to radioimmunoassay of cAMP in tissue sample, it is essential to remove protein from sample, because plasma itself binds 3 H-cAMP (in our unpublished data) and contains phosphodiesterase which destroys cAMP. The TCA-ether,^{1),4),5)} PCA-KOH, and the Somogi methods have been employed for this purpose. According to our experience, the Somogi method was the most suitable at least for our radioimmunoassay, because the other two methods were troublesome and time-consuming. In the case of TCA-ether method, it was essential for the sample to be dried up and be redissolved in buffer prior to the assay, and the TCA-ether method without the above procedures often resulted in indefinite data.

Acknowledgement

The authors wish to express our thanks to Dr. S. Makino, Department of Medicine, Dokkyo Medical College, for suggesting this investigation as well as for constant guidance in the course of the work.

References

- 1) A.L. Steiner, D.M. Kipnis, R. Utiger and C. Parker: *Proc. Natl. Acad. Sci.*

- U. S., 64, 367 (1969)
- 2) A.L. Steiner, C.W. Parker and D.M. Kipnis: *J. Biol. Chem.*, 247, 1106 (1972)
- 3) I. Weinryb, I.M. Michel and S.M. Hess: *Anal. Biochem.*, 45, 659 (1972)
- 4) T. Okabayashi, S. Mihara, M. Nakamura, A. Tanaka and F. Sagara: *Amino Acid and Nucleic Acid*, 24, 28 (1971)
- 5) A.G. Gilman: *Proc. Natl. Acad. Sci. U. S.*, 67, 305 (1970)
- 6) B.L. Brown, J.D.M. Albano, R.P. Ekins and A.M. Scherzi: *Biochem. J.*, 121, 561 (1971)
- 7) A.L. Latner and K. Prudhoe: *Clinica Chimica Acta*, 48, 353 (1973)
- 8) T. Okabayashi: *Hormon to Rinsho*, 21, 287 (1973)
- 9) H.L. Cailla, G.S. Gros, E.J.P. Jolu, M. A. Delaage and R.C. Depieds: *Anal. Biochem.*, 56, 383 (1973)

要 旨

Dextran-coated charcoal を用いた簡便なアデノシン-3':5'-リン酸 のラジオイムノアッセイ

加藤良樹, 竹沢順一, 鈴木守郎, 国中 明, 吉野 広

ヤマサ醤油研究所

千葉県銚子市新生2-550

cAMP のラジオイムノアッセイにおいて、従来使用されていなかった dextran-coated charcoal (DCC) が一定の条件下で、遊離 cAMP (F) と抗体結合 cAMP (B) の分離に有効なことを知り、これを利用して簡便かつ経済的なアッセイ方法を確立した。すなわち反応液への DCC 添加接触後、4 分間以内では F のみが定量的に吸着され、B はほとんど吸着されないの、この条件下で DCC による F と B の分離を定量的におこなうことに成功したのである。この方法で得られたデータはセルロースエステルフィルター分離法で行なった他の実験データとよく一致した。この方法により cAMP のイムノアッセイを binding assay と同じように、簡便かつ安価に行なうことが可能になった。