Erratum

The Validity of Mass Body Temperature Screening with Ear Thermometers in a Warm Thermal Environment

Tatsuhiko Suzuki,¹ Koji Wada,² Yuko Wada,¹ Hideaki Kagitani,¹ Tetsuya Arioka,¹
Koji Maeda¹ and Kenichi Kida¹

¹Terumo Corporation, Tokyo, Japan

²Department of Preventive Medicine and Public Health, Kitasato University School of Medicine, Kanagawa, Japan

Tohoku J. Exp. Med., 222: 89-95, 2010.

In the version of this article published in the October issue, 2010, the graphs shown in (a) and (b) of Figure 2 are adversely displayed (page 92). The corrected figure is shown below. The data in each graph remain unchanged.

(a) Cold environment

(b) Warm environment

Fig. 2. Distribution of ear and facial temperature plotted against axillary temperature in cold and warm environments.

(a) Cold environment (mean ambient temperature was 12.6°C, mean humidity was 33.0%, and mean wind speed was 1.3 m/s). (b) Warm environment (mean ambient temperature was 20.0°C, mean humidity was 35.0%, and mean wind speed was 0.0 m/s). □: Ear temperature; x: facial temperature.